

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD022121\
 Data File : PD061333.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Feb 2021 14:47
 Operator : DD\AJ
 Sample : INDB301
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 22 03:50:35 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD022121CLP.M
 Quant Title : GC Extractables
 QLast Update : Mon Feb 22 03:49:37 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.487	3.982	17519815	163.6E6	20.000	20.000
27)	SA Decachlor...	8.261	9.036	48463966	284.7E6	40.000	40.000
Target Compounds							
5)	MB Aldrin	4.766	5.467	24563847	219.7E6	20.000	20.000
6)	B beta-BHC	4.462	4.824	12304984	100.9E6	20.000	20.000
7)	B delta-BHC	4.668	5.040	25604174	229.1E6	20.000	20.000
8)	B Heptachlo...	5.210	5.853	24002856	197.4E6	20.000	20.000
10)	B trans-Chl...	5.437	6.087	24699766	202.2E6	20.000	20.000
11)	B cis-Chlor...	5.496	6.159	24325321	192.8E6	20.000	20.000
12)	B 4,4'-DDE	5.660	6.316	46784447	386.3E6	40.000	40.000
15)	B Endosulfa...	6.326	6.893	41068000	305.8E6	40.000	40.000
18)	B Endrin al...	6.495	7.016	35760224	249.2E6	40.000	40.000
19)	B Endosulfa...	6.709	7.239	41742403	294.7E6	40.000	40.000
21)	B Endrin ke...	7.202	7.709	47993042	290.7E6	40.000	40.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD022121\
Data File : PD061333.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Feb 2021 14:47
Operator : DD\AJ
Sample : INDB301
Misc :
ALS Vial : 14 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 22 03:50:35 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD022121CLP.M
Quant Title : GC Extractables
QLast Update : Mon Feb 22 03:49:37 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm




Instrument :
ECD_D
ClientSampled :



#1 Tetrachloro-m-xylene


R.T.: 3.487 min
Delta R.T.: 0.000 min
Response: 17519815
Conc: 20.00 ng/ml

Instrument :
ECD_D
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.982 min
Delta R.T.: 0.000 min
Response: 163639607
Conc: 20.00 ng/ml




#5 Aldrin

R.T.: 4.766 min
Delta R.T.: 0.000 min
Response: 24563847
Conc: 20.00 ng/ml



#5 Aldrin


R.T.: 5.467 min
Delta R.T.: 0.000 min
Response: 219721365
Conc: 20.00 ng/ml



#6 beta-BHC


R.T.: 4.462 min
Delta R.T.: 0.000 min
Response: 12304984
Conc: 20.00 ng/ml

Instrument :
ECD_D
ClientSampleId :




#6 beta-BHC

R.T.: 4.824 min
Delta R.T.: 0.000 min
Response: 100945677
Conc: 20.00 ng/ml




#7 delta-BHC

R.T.: 4.668 min
Delta R.T.: 0.000 min
Response: 25604174
Conc: 20.00 ng/ml



#7 delta-BHC


R.T.: 5.040 min
Delta R.T.: 0.000 min
Response: 229081240
Conc: 20.00 ng/ml



#8 Heptachlor epoxide


R.T.: 5.210 min
Delta R.T.: 0.000 min
Response: 24002856
Conc: 20.00 ng/ml

Instrument :
ECD_D
ClientSampleId :




#8 Heptachlor epoxide

R.T.: 5.853 min
Delta R.T.: 0.000 min
Response: 197366704
Conc: 20.00 ng/ml




#10 trans-Chlordane

R.T.: 5.437 min
Delta R.T.: 0.000 min
Response: 24699766
Conc: 20.00 ng/ml



#10 trans-Chlordane


R.T.: 6.087 min
Delta R.T.: 0.000 min
Response: 202243599
Conc: 20.00 ng/ml



#11 cis-Chlordane


R.T.: 5.496 min
Delta R.T.: 0.000 min
Response: 24325321
Conc: 20.00 ng/ml

Instrument :
ECD_D
ClientSampleId :




#11 cis-Chlordane

R.T.: 6.159 min
Delta R.T.: 0.000 min
Response: 192789335
Conc: 20.00 ng/ml




#12 4,4'-DDE

R.T.: 5.660 min
Delta R.T.: 0.000 min
Response: 46784447
Conc: 40.00 ng/ml



#12 4,4'-DDE


R.T.: 6.316 min
Delta R.T.: 0.000 min
Response: 386348645
Conc: 40.00 ng/ml



#15 Endosulfan II


R.T.: 6.326 min
Delta R.T.: 0.000 min
Response: 41068000
Conc: 40.00 ng/ml

Instrument :
ECD_D
ClientSampleId :




#15 Endosulfan II

R.T.: 6.893 min
Delta R.T.: 0.000 min
Response: 305798050
Conc: 40.00 ng/ml




#18 Endrin aldehyde

R.T.: 6.495 min
Delta R.T.: 0.000 min
Response: 35760224
Conc: 40.00 ng/ml



#18 Endrin aldehyde


R.T.: 7.016 min
Delta R.T.: 0.000 min
Response: 249176569
Conc: 40.00 ng/ml



#19 Endosulfan Sulfate


R.T.: 6.709 min
Delta R.T.: 0.000 min
Response: 41742403
Conc: 40.00 ng/ml

Instrument :
ECD_D
ClientSampleId :




#19 Endosulfan Sulfate

R.T.: 7.239 min
Delta R.T.: 0.000 min
Response: 294715234
Conc: 40.00 ng/ml




#21 Endrin ketone

R.T.: 7.202 min
Delta R.T.: 0.000 min
Response: 47993042
Conc: 40.00 ng/ml



#21 Endrin ketone

R.T.: 7.709 min
Delta R.T.: 0.000 min
Response: 290735414
Conc: 40.00 ng/ml



#27 Decachlorobiphenyl


R.T.: 8.261 min

Delta R.T.: 0.000 min

Response: 48463966

Conc: 40.00 ng/ml

Instrument :
ECD_D
ClientSampleId :



#27 Decachlorobiphenyl

R.T.: 9.036 min

Delta R.T.: 0.000 min

Response: 284701272

Conc: 40.00 ng/ml