

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD022121\
 Data File : PD061374.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Feb 2021 03:11
 Operator : DD\AJ
 Sample : M1412-20
 Misc :
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 22 05:09:28 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD022121CLP.M
 Quant Title : GC Extractables
 QLast Update : Mon Feb 22 04:35:05 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.485	3.982	9156860	72999174	9.944m	8.730
27)	SA Decachlor...	8.261	9.037	44623061	102.6E6	36.269m	14.275m#
Target Compounds							
2)	A alpha-BHC	3.921	4.368	1385834	13978410	1.048m	1.089
3)	MA gamma-BHC...	4.205	4.652	6487935	9619816	5.043m	0.823 #
4)	MA Heptachlor	4.519	5.161	1690902	22633916	1.281m	1.916 #
7)	B delta-BHC	4.670	5.040	3263688	34182258	2.525m	2.930m
9)	A Endosulfan I	5.556	6.214	4336250	57015352	3.676	6.173m#
10)	B trans-Chl...	5.437	6.092	4470654	57615660	3.583m	5.771m#
13)	MA Dieldrin	5.784	6.477	11420513	28643645	8.941m	2.914 #
16)	A 4,4'-DDD	6.178	6.804	2511132	70523262	2.473m	9.203m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD022121\
Data File : PD061374.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Feb 2021 03:11
Operator : DD\AJ
Sample : M1412-20
Misc :
ALS Vial : 51 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 22 05:09:28 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD022121CLP.M
Quant Title : GC Extractables
QLast Update : Mon Feb 22 04:35:05 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

