

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Data\PD022121\  
 Data File : PD061334.D  
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
 Acq On : 21 Feb 2021 15:01  
 Operator : DD\AJ  
 Sample : INDA401  
 Misc :  
 ALS Vial : 15 Sample Multiplier: 1

Instrument :  
 ECD\_D  
 ClientSampleId :  
 INDA401

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Feb 22 04:23:23 2021  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Method\PD022121CLP.M  
 Quant Title : GC Extractables  
 QLast Update : Mon Feb 22 04:21:33 2021  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 1 µl  
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1  
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm


|                             | Compound        | RT#1  | RT#2  | Resp#1   | Resp#2   | ng/ml   | ng/ml   |
|-----------------------------|-----------------|-------|-------|----------|----------|---------|---------|
| -----                       |                 |       |       |          |          |         |         |
| System Monitoring Compounds |                 |       |       |          |          |         |         |
| 1)                          | SA Tetrachlo... | 3.487 | 3.982 | 35844899 | 330.4E6  | 38.914  | 38.645  |
| 27)                         | SA Decachlor... | 8.260 | 9.036 | 92481954 | 546.3E6  | 75.523  | 75.277  |
| Target Compounds            |                 |       |       |          |          |         |         |
| 2)                          | A alpha-BHC     | 3.927 | 4.370 | 54070658 | 509.1E6  | 41.096  | 39.033  |
| 3)                          | MA gamma-BHC... | 4.211 | 4.655 | 51654710 | 464.2E6  | 40.226  | 39.233  |
| 4)                          | MA Heptachlor   | 4.515 | 5.163 | 52356367 | 461.5E6  | 39.735  | 38.781  |
| 9)                          | A Endosulfan I  | 5.550 | 6.210 | 45681782 | 363.1E6  | 39.095  | 39.041  |
| 13)                         | MA Dieldrin     | 5.794 | 6.466 | 103.2E6  | 774.2E6  | 80.729  | 78.064  |
| 14)                         | MA Endrin       | 6.052 | 6.684 | 83311794 | 601.5E6  | 79.256  | 76.947  |
| 16)                         | A 4,4'-DDD      | 6.176 | 6.805 | 82331974 | 604.2E6  | 80.764  | 77.957  |
| 17)                         | MA 4,4'-DDT     | 6.418 | 7.104 | 83252435 | 579.4E6  | 82.647  | 77.765  |
| 20)                         | A Methoxychlor  | 6.965 | 7.561 | 216.8E6  | 1333.8E6 | 387.073 | 379.317 |
| -----                       |                 |       |       |          |          |         |         |

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Data\PD022121\  
Data File : PD061334.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 21 Feb 2021 15:01  
Operator : DD\AJ  
Sample : INDA401  
Misc :  
ALS Vial : 15 Sample Multiplier: 1

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Feb 22 04:23:23 2021  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Method\PD022121CLP.M  
Quant Title : GC Extractables  
QLast Update : Mon Feb 22 04:21:33 2021  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1 µl  
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1  
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



Instrument :  
ECD\_D  
ClientSample :  
INDA401