

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD022219CLP.M
 Title : GC Extractables
 Last Update : Fri Feb 22 03:43:16 2019
 Response Via : Initial Calibration

Calibration Files

3 =PD051315.D 1 =PD051313.D 2 =PD051314.D
 4 =PD051316.D 5 =PD051317.D

	Compound	3	1	2	4	5	Avg	%RSD
1) SA	Tetrachloro-m-xylene	1.522	1.492	1.516	1.526	1.513	1.514 E6	0.88
2) A	alpha-BHC						0.000	-1.00
3) MA	gamma-BHC (Lindane)						0.000	-1.00
4) MA	Heptachlor						0.000	-1.00
5) MB	Aldrin						0.000	-1.00
6) B	beta-BHC						0.000	-1.00
7) B	delta-BHC						0.000	-1.00
8) B	Heptachlor epoxide						0.000	-1.00
9) A	Endosulfan I						0.000	-1.00
10) B	trans-Chlordane						0.000	-1.00
11) B	cis-Chlordane						0.000	-1.00
12) B	4,4'-DDE						0.000	-1.00
13) MA	Dieldrin						0.000	-1.00
14) MA	Endrin						0.000	-1.00
15) B	Endosulfan II						0.000	-1.00
16) A	4,4'-DDD						0.000	-1.00
17) MA	4,4'-DDT						0.000	-1.00
18) B	Endrin aldehyde						0.000	-1.00
19) B	Endosulfan Sulfate						0.000	-1.00
20) A	Methoxychlor						0.000	-1.00
21) B	Endrin ketone						0.000	-1.00
22)	Toxaphene-1	1.363	1.332	1.320	1.400	1.403	1.364 E4	2.79
23)	Toxaphene-2	2.292	2.197	2.212	2.320	2.417	2.288 E4	3.90
24)	Toxaphene-3	3.853	3.465	3.596	4.014	4.097	3.805 E4	7.08
25)	Toxaphene-4	3.870	3.454	3.694	4.093	4.194	3.861 E4	7.76
26)	Toxaphene-5	4.385	4.111	4.224	4.560	4.536	4.363 E4	4.47
27) SA	Decachlorobiphenyl	1.257	1.431	1.330	1.191	1.124	1.267 E6	9.46

Signal #2 Calibration Files

3 =PD051315.D 1 =PD051313.D 2 =PD051314.D
 4 =PD051316.D 5 =PD051317.D

	Compound	3	1	2	4	5	Avg	%RSD
1) SA	Tetrachloro-m-xylene	3.296	3.616	3.479	3.087	2.819	3.259 E7	9.71
2) A	alpha-BHC						0.000	-1.00
3) MA	gamma-BHC (Lindane)						0.000	-1.00
4) MA	Heptachlor						0.000	-1.00
5) MB	Aldrin						0.000	-1.00
6) B	beta-BHC						0.000	-1.00
7) B	delta-BHC						0.000	-1.00
8) B	Heptachlor epoxide						0.000	-1.00
9) A	Endosulfan I						0.000	-1.00
10) B	trans-Chlordane						0.000	-1.00
11) B	cis-Chlordane						0.000	-1.00
12) B	4,4'-DDE						0.000	-1.00
13) MA	Dieldrin						0.000	-1.00
14) MA	Endrin						0.000	-1.00
15) B	Endosulfan II						0.000	-1.00
16) A	4,4'-DDD						0.000	-1.00
17) MA	4,4'-DDT						0.000	-1.00
18) B	Endrin aldehyde						0.000	-1.00
19) B	Endosulfan Sulfate						0.000	-1.00
20) A	Methoxychlor						0.000	-1.00
21) B	Endrin ketone						0.000	-1.00
22)	Toxaphene-1	1.463	1.681	1.599	1.446	1.325	1.503 E5	9.25
23)	Toxaphene-2	2.925	3.189	3.046	2.736	2.592	2.898 E5	8.23

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
Method File : PD022219CLP.M
Title : GC Extractables
Last Update : Fri Feb 22 03:43:16 2019
Response Via : Initial Calibration

Calibration Files

3 =PD051315.D 1 =PD051313.D 2 =PD051314.D
4 =PD051316.D 5 =PD051317.D

	Compound	3	1	2	4	5	Avg		%RSD
24)	Toxaphene-3	2.413	2.702	2.548	2.313	2.179	2.431	E5	8.36
25)	Toxaphene-4	6.708	7.329	6.991	6.458	6.093	6.716	E5	7.09
26)	Toxaphene-5	2.899	3.124	3.033	2.758	2.577	2.878	E5	7.57
27) SA	Decachlorobiphenyl	1.963	2.296	2.132	1.788	1.617	1.959	E7	13.72

(#) = Out of Range