

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD022219CLP.M
 Title : GC Extractables
 Last Update : Fri Feb 22 04:30:42 2019
 Response Via : Initial Calibration

Calibration Files

3 =PD051323.D 1 =PD051319.D 2 =PD051321.D
 4 =PD051325.D 5 =PD051327.D

	Compound	3	1	2	4	5	Avg	%RSD
1) SA	Tetrachloro-m-xylene	1.390	1.395	1.423	1.416	1.419	1.409 E6	1.05
2) A	alpha-BHC						0.000	-1.00
3) MA	gamma-BHC (Lindane)						0.000	-1.00
4) MA	Heptachlor						0.000	-1.00
5) MB	Aldrin	2.109	2.010	2.077	2.238	2.300	2.147 E6	5.55
6) B	beta-BHC	9.541	9.596	9.927	9.479	9.262	9.561 E5	2.52
7) B	delta-BHC	2.207	2.078	2.171	2.357	2.424	2.247 E6	6.25
8) B	Heptachlor epoxide	2.073	2.066	2.085	2.153	2.164	2.108 E6	2.21
9) A	Endosulfan I						0.000	-1.00
10) B	trans-Chlordane	2.142	2.095	2.138	2.246	2.281	2.180 E6	3.63
11) B	cis-Chlordane	2.105	2.088	2.119	2.183	2.192	2.137 E6	2.21
12) B	4,4'-DDE	2.138	1.911	2.035	2.302	2.372	2.151 E6	8.78
13) MA	Dieldrin						0.000	-1.00
14) MA	Endrin						0.000	-1.00
15) B	Endosulfan II	2.058	1.976	2.044	2.138	2.146	2.072 E6	3.42
16) A	4,4'-DDD						0.000	-1.00
17) MA	4,4'-DDT						0.000	-1.00
18) B	Endrin aldehyde	1.623	1.630	1.642	1.657	1.646	1.640 E6	0.82
19) B	Endosulfan Sulfate	1.824	1.857	1.860	1.854	1.839	1.847 E6	0.81
20) A	Methoxychlor						0.000	-1.00
21) B	Endrin ketone	2.015	2.020	2.032	2.049	2.040	2.031 E6	0.70
22)	Toxaphene-1						0.000	-1.00
23)	Toxaphene-2						0.000	-1.00
24)	Toxaphene-3						0.000	-1.00
25)	Toxaphene-4						0.000	-1.00
26)	Toxaphene-5						0.000	-1.00
27) SA	Decachlorobiphenyl	1.174	1.353	1.272	1.119	1.059	1.196 E6	9.85

Signal #2 Calibration Files

3 =PD051323.D 1 =PD051319.D 2 =PD051321.D
 4 =PD051325.D 5 =PD051327.D

	Compound	3	1	2	4	5	Avg	%RSD
1) SA	Tetrachloro-m-xylene	3.014	3.339	3.247	2.844	2.644	3.018 E7	9.46
2) A	alpha-BHC						0.000	-1.00
3) MA	gamma-BHC (Lindane)						0.000	-1.00
4) MA	Heptachlor						0.000	-1.00
5) MB	Aldrin	3.955	4.351	4.235	3.765	3.484	3.958 E7	8.87
6) B	beta-BHC	1.974	2.299	2.166	1.816	1.665	1.984 E7	12.91
7) B	delta-BHC	3.985	4.226	4.407	3.972	3.764	4.071 E7	6.12
8) B	Heptachlor epoxide	3.549	4.039	3.875	3.364	3.076	3.580 E7	10.81
9) A	Endosulfan I						0.000	-1.00
10) B	trans-Chlordane	3.617	3.977	3.870	3.418	3.182	3.613 E7	9.00
11) B	cis-Chlordane	3.477	3.823	3.712	3.267	3.030	3.462 E7	9.33
12) B	4,4'-DDE	3.264	3.557	3.452	3.127	2.937	3.267 E7	7.61
13) MA	Dieldrin						0.000	-1.00
14) MA	Endrin						0.000	-1.00
15) B	Endosulfan II	2.750	3.073	2.955	2.574	2.372	2.745 E7	10.30
16) A	4,4'-DDD						0.000	-1.00
17) MA	4,4'-DDT						0.000	-1.00
18) B	Endrin aldehyde	2.252	2.551	2.437	2.091	1.907	2.247 E7	11.54
19) B	Endosulfan Sulfate	2.533	2.905	2.743	2.354	2.158	2.539 E7	11.74
20) A	Methoxychlor						0.000	-1.00
21) B	Endrin ketone	2.393	2.632	2.538	2.268	2.100	2.386 E7	8.89
22)	Toxaphene-1						0.000	-1.00
23)	Toxaphene-2						0.000	-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
Method File : PD022219CLP.M
Title : GC Extractables
Last Update : Fri Feb 22 04:30:42 2019
Response Via : Initial Calibration

Calibration Files

3 =PD051323.D 1 =PD051319.D 2 =PD051321.D
4 =PD051325.D 5 =PD051327.D

	Compound	3	1	2	4	5	Avg	%RSD
24)	Toxaphene-3						0.000	-1.00
25)	Toxaphene-4						0.000	-1.00
26)	Toxaphene-5						0.000	-1.00
27) SA	Decachlorobiphenyl	1.884	2.201	2.079	1.721	1.550	1.887 E7	13.93

(#) = Out of Range