

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : DTX022319.M
 Title : GC Extractables
 Last Update : Fri Feb 22 06:25:42 2019
 Response Via : Initial Calibration

Calibration Files

500 =PD051346.D 1000 =PD051344.D 750 =PD051345.D
 250 =PD051347.D 50 =PD051348.D

	Compound	500	1000	750	250	50	Avg		%RSD

1) SA	Tetrachloro-m-xylene	1.416	1.474	1.413	1.352	1.435	1.418	E6	3.11
2)	Toxaphene-1	1.176	0.986	1.168	1.264	1.231	1.165	E4	9.22
3)	Toxaphene-2	3.137	3.234	3.110	3.191	3.301	3.195	E4	2.40
4)	Toxaphene-3	1.632	1.492	1.495	1.642	1.385	1.529	E4	7.06
5)	Toxaphene-4	2.397	2.382	2.331	2.358	2.089	2.311	E4	5.49
6)	Toxaphene-5	3.614	3.691	3.602	3.768	3.691	3.673	E4	1.83
7) SA	Decachlorobiphenyl	1.049	1.012	1.007	1.106	1.358	1.106	E6	13.22

Signal #2 Calibration Files

500 =PD051346.D 1000 =PD051344.D 750 =PD051345.D
 250 =PD051347.D 50 =PD051348.D

	Compound	500	1000	750	250	50	Avg		%RSD

1) SA	Tetrachloro-m-xylene	2.709	2.683	2.722	2.887	3.482	2.897	E7	11.63
2)	Toxaphene-1	7.925	7.384	7.779	8.471	9.610	8.234	E4	10.47
3)	Toxaphene-2	1.809	1.751	1.698	1.776	2.024	1.812	E5	6.93
4)	Toxaphene-3	6.611	6.217	6.528	6.930	7.781	6.813	E5	8.77
5)	Toxaphene-4	4.002	3.880	3.978	4.245	4.463	4.114	E5	5.76
6)	Toxaphene-5	1.800	1.749	1.737	1.822	1.768	1.775	E5	2.00
7) SA	Decachlorobiphenyl	1.786	1.673	1.695	1.948	2.488	1.918	E7	17.54

(#) = Out of Range