

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD022422\
 Data File : PD068165.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Feb 2022 09:34
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 25 01:01:12 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD022222.M
 Quant Title : GC Extractables
 QLast Update : Tue Feb 22 13:16:10 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.449	4.176	74225128	455.9E6	50.986	49.297
28)	SA Decachlor...	8.213	9.369	68538708	254.5E6	47.829	51.121
Target Compounds							
2)	A alpha-BHC	3.889	4.571	116.4E6	667.5E6	52.351	50.495
3)	MA gamma-BHC...	4.172	4.862	105.9E6	596.9E6	51.818	50.271
4)	MA Heptachlor	4.472	5.389	100.1E6	517.4E6	51.538	50.618
5)	MB Aldrin	4.720	5.702	110.2E6	575.0E6	50.941	49.968
6)	B beta-BHC	4.422	5.024	42821920	255.3E6	50.261	49.339
7)	B delta-BHC	4.627	5.250	105.9E6	593.9E6	51.647	50.176
8)	B Heptachlo...	5.164	6.090	97761225	503.3E6	50.156	49.849
9)	A Endosulfan I	5.503	6.452	90094830	449.1E6	50.254	50.085
10)	B gamma-Chl...	5.391	6.326	99814373	496.1E6	50.186	50.572
11)	B alpha-Chl...	5.450	6.401	98042005	478.9E6	50.139	49.974
12)	B 4,4'-DDE	5.616	6.551	93036348	451.4E6	51.074	49.971
13)	MA Dieldrin	5.747	6.710	100.3E6	480.4E6	50.566	49.953
14)	MA Endrin	6.005	6.931	81776349	342.7E6	49.841	47.819
15)	B Endosulfa...	6.279	7.135	85113138	379.6E6	50.632	49.711
16)	A 4,4'-DDD	6.133	7.041	77458024	348.1E6	50.533	49.668
17)	MA 4,4'-DDT	6.374	7.345	72778196	292.9E6	52.900	51.663
18)	B Endrin al...	6.449	7.260	56549714	262.2E6	51.227	50.639
19)	B Endosulfa...	6.661	7.484	80553284	335.9E6	51.235	49.157
20)	A Methoxychlor	6.926	7.798	40657960	143.4E6	52.256	52.984
21)	B Endrin ke...	7.155	7.959	91649175	367.3E6	51.294	51.102
22)	Mirex	7.341	8.426	70430964	250.1E6	49.277	49.897

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD022422\
Data File : PD068165.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Feb 2022 09:34
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 25 01:01:12 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD022222.M
Quant Title : GC Extractables
QLast Update : Tue Feb 22 13:16:10 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

