

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD022422\
 Data File : PD068187.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Feb 2022 18:24
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 25 06:19:57 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD022222.M
 Quant Title : GC Extractables
 QLast Update : Tue Feb 22 13:16:10 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.450	4.176	73394513	450.3E6	50.416	48.689
2)	SA Decachlor...	8.213	9.370	72138059	259.1E6	50.340	52.055
Target Compounds							
2)	A alpha-BHC	3.889	4.571	116.3E6	664.1E6	52.305	50.237
3)	MA gamma-BHC...	4.173	4.862	106.2E6	596.2E6	51.985	50.212
4)	MA Heptachlor	4.472	5.390	101.2E6	529.3E6	52.151	51.781
5)	MB Aldrin	4.720	5.703	111.3E6	578.3E6	51.461	50.250
6)	B beta-BHC	4.422	5.024	43187401	255.9E6	50.690	49.453
7)	B delta-BHC	4.627	5.250	108.2E6	601.0E6	52.771	50.772
8)	B Heptachlo...	5.164	6.091	99541324	511.9E6	51.069	50.706
9)	A Endosulfan I	5.503	6.453	91750838	451.5E6	51.178	50.361
10)	B gamma-Chl...	5.391	6.327	101.8E6	500.5E6	51.192	51.015
11)	B alpha-Chl...	5.450	6.401	99731246	483.8E6	51.003	50.493
12)	B 4,4'-DDE	5.615	6.552	95684066	456.4E6	52.528	50.518
13)	MA Dieldrin	5.747	6.711	102.8E6	490.3E6	51.851	50.978
14)	MA Endrin	6.005	6.931	83583406	345.3E6	50.942	48.176
15)	B Endosulfa...	6.279	7.136	87141273	392.1E6	51.838	51.342
16)	A 4,4'-DDD	6.133	7.042	80128406	364.4E6	52.275	51.989
17)	MA 4,4'-DDT	6.374	7.345	73632010	308.2E6	53.521	54.362
18)	B Endrin al...	6.449	7.260	57364123	270.4E6	51.964	52.216
19)	B Endosulfa...	6.662	7.484	82483572	359.6E6	52.463	52.629
20)	A Methoxychlor	6.925	7.798	40783865	151.3E6	52.418	55.886
21)	B Endrin ke...	7.154	7.960	94922646	379.9E6	53.126	52.848
22)	Mirex	7.340	8.426	72701581	260.1E6	50.866	51.889

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD022422\
Data File : PD068187.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Feb 2022 18:24
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 25 06:19:57 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD022222.M
Quant Title : GC Extractables
QLast Update : Tue Feb 22 13:16:10 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

