

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD022624\
 Data File : PD081643.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Feb 2024 12:21
 Operator : AR\AJ
 Sample : PB159203BL
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PBLK203

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 26 21:24:46 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD021224CLP.M
 Quant Title : GC Extractables
 QLast Update : Mon Feb 12 17:06:05 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.572	2.790	31909356	62752996	23.340	25.597
27)	SA Decachlor...	9.122	7.930	51741038	75530856	26.818	30.447
Target Compounds							
3)	MA gamma-BHC...	0.000	3.633	0	113880	N.D.	<MDL
4)	MA Heptachlor	0.000	3.953	0	36313	N.D.	<MDL
5)	MB Aldrin	0.000	4.270	0	110083	N.D.	<MDL
6)	B beta-BHC	0.000	3.923	0	119998	N.D.	<MDL
18)	B Endrin al...	0.000	6.138	0	488240	N.D.	0.225


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD022624\
Data File : PD081643.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Feb 2024 12:21
Operator : AR\AJ
Sample : PB159203BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 26 21:24:46 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD021224CLP.M
Quant Title : GC Extractables
QLast Update : Mon Feb 12 17:06:05 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.572 min
Delta R.T.: 0.006 min
Response: 31909356
Conc: 23.34 ng/ml

Instrument :
ECD_D
ClientSampleId :
PBLK203




#1 Tetrachloro-m-xylene

R.T.: 2.790 min
Delta R.T.: -0.004 min
Response: 62752996
Conc: 25.60 ng/ml




#18 Endrin aldehyde

R.T.: 0.000 min
Exp R.T. : 6.960 min
Response: 0
Conc: N.D.



#18 Endrin aldehyde


R.T.: 6.138 min
Delta R.T.: 0.007 min
Response: 488240
Conc: 0.23 ng/ml



#27 Decachlorobiphenyl

R.T.: 9.122 min
Delta R.T.: 0.008 min
Response: 51741038
Conc: 26.82 ng/ml

Instrument :
ECD_D
ClientSampleId :
PBLK203



#27 Decachlorobiphenyl

R.T.: 7.930 min
Delta R.T.: -0.003 min
Response: 75530856
Conc: 30.45 ng/ml