

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : DTX030122.M
 Title : GC Extractables
 Last Update : Tue Mar 01 07:11:37 2022
 Response Via : Initial Calibration

Calibration Files

500 =PD068264.D 1000 =PD068262.D 750 =PD068263.D
 250 =PD068265.D 100 =PD068266.D

	Compound	500	1000	750	250	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	1.321	1.259	1.287	1.265	1.186	1.264	E6	3.92
2)	Toxaphene-1	1.104	1.011	1.046	1.142	1.015	1.064	E4	5.41
3)	Toxaphene-2	1.664	1.355	1.568	1.726	1.479	1.558	E4	9.48
4)	Toxaphene-3	2.622	2.144	2.549	2.702	2.481	2.499	E4	8.61
5)	Toxaphene-4	2.933	2.649	2.754	2.843	2.709	2.778	E4	4.04
6)	Toxaphene-5	2.921	2.464	2.727	2.908	2.634	2.731	E4	7.04
7) SA	Decachlorobiphenyl	1.325	1.180	1.207	1.292	1.229	1.247	E6	4.83

Signal #2 Calibration Files

500 =PD068264.D 1000 =PD068262.D 750 =PD068263.D
 250 =PD068265.D 100 =PD068266.D

	Compound	500	1000	750	250	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	1.008	0.933	0.951	1.022	1.042	0.991	E7	4.75
2)	Toxaphene-1	3.088	2.670	2.771	2.974	2.903	2.881	E4	5.71
3)	Toxaphene-2	1.215	1.085	1.122	1.227	1.231	1.176	E5	5.76
4)	Toxaphene-3	1.509	1.387	1.408	1.521	1.554	1.476	E5	5.00
5)	Toxaphene-4	6.010	5.643	5.549	5.924	6.197	5.865	E4	4.54
6)	Toxaphene-5	1.025	0.936	0.982	1.040	1.010	0.999	E5	4.09
7) SA	Decachlorobiphenyl	6.199	5.614	5.744	6.314	6.401	6.054	E6	5.83

(#) = Out of Range