

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD030122\
 Data File : PD068252.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Feb 2022 17:56
 Operator : AR\AJ
 Sample : PSTDICC100
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDICC100

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 01 06:19:33 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD030122.M
 Quant Title : GC Extractables
 QLast Update : Tue Mar 01 06:18:38 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.450	4.176	116.1E6	887.4E6	98.376	94.341
28)	SA Decachlor...	8.213	9.374	110.1E6	540.7E6	96.313	92.867
Target Compounds							
2)	A alpha-BHC	3.890	4.572	184.8E6	1341.7E6	103.332	97.699
3)	MA gamma-BHC...	4.173	4.863	169.2E6	1200.0E6	102.630	97.082
4)	MA Heptachlor	4.472	5.390	157.0E6	1086.9E6	101.762	96.452
5)	MB Aldrin	4.720	5.703	163.6E6	1128.0E6	101.157	96.026
6)	B beta-BHC	4.423	5.025	67921224	496.8E6	96.514	93.668
7)	B delta-BHC	4.628	5.251	172.1E6	1206.6E6	102.713	97.441
8)	B Heptachlo...	5.165	6.091	149.2E6	1003.1E6	100.870	95.375
9)	A Endosulfan I	5.504	6.454	137.8E6	908.6E6	99.976	94.981
10)	B gamma-Chl...	5.391	6.328	151.7E6	1001.3E6	100.870	95.956
11)	B alpha-Chl...	5.450	6.403	148.3E6	971.4E6	100.105	95.398
12)	B 4,4'-DDE	5.617	6.553	143.3E6	941.9E6	102.855	96.479
13)	MA Dieldrin	5.748	6.713	152.8E6	986.2E6	101.985	96.143
14)	MA Endrin	6.006	6.933	128.7E6	749.6E6	102.311	96.978
15)	B Endosulfa...	6.279	7.137	128.6E6	778.0E6	99.731	95.744
16)	A 4,4'-DDD	6.134	7.043	122.7E6	751.4E6	101.214	96.371
17)	MA 4,4'-DDT	6.374	7.346	118.5E6	699.7E6	103.321	97.893
18)	B Endrin al...	6.450	7.262	89220493	558.5E6	97.166	94.118
19)	B Endosulfa...	6.662	7.486	125.8E6	734.8E6	99.267	96.649
20)	A Methoxychlor	6.925	7.799	64080923	331.5E6	97.409	94.925
21)	B Endrin ke...	7.155	7.961	148.1E6	787.2E6	99.778	94.963
22)	Mirex	7.340	8.428	119.1E6	583.0E6	96.263	93.301

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD030122\
Data File : PD068252.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Feb 2022 17:56
Operator : AR\AJ
Sample : PSTDICC100
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDICC100

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 01 06:19:33 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD030122.M
Quant Title : GC Extractables
QLast Update : Tue Mar 01 06:18:38 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

