

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD030219\
 Data File : PD051515.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 02 Mar 2019 2:26
 Operator : AJ\SJ
 Sample : PTOXCCC500
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PTOXCCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 02 03:59:14 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX030219.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 01 23:57:16 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.224	4.044	70263619	1452.7E6	52.015	50.179
7) SA Decachlor...	7.824	9.156	60961617	910.0E6	52.172	48.562
Target Compounds						
2) Toxaphene-1	5.524	6.382	6114295	45814147	486.717	493.915
3) Toxaphene-2	5.956	6.739	17447437	99995891	525.198	503.696
4) Toxaphene-3	6.329	7.224	8954776	365.0E6	558.878	513.758
5) Toxaphene-4	6.541	7.310	13558279	225.1E6	545.667	529.266
6) Toxaphene-5	6.668	7.451	21373971	95826550	542.696	532.793

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD030219\
Data File : PD051515.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 02 Mar 2019 2:26
Operator : AJ\SJ
Sample : PTOXCCC500
Misc :
ALS Vial : 17 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 02 03:59:14 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX030219.M
Quant Title : GC Extractables
QLast Update : Fri Mar 01 23:57:16 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

