

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD030323\
 Data File : PD074110.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Mar 2023 11:50
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 02 21:29:07 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD030123.M
 Quant Title : GC Extractables
 QLast Update : Wed Mar 01 01:35:06 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.556	2.774	106.2E6	61265617	46.354	51.726
28)	SA Decachlor...	9.105	7.931	136.1E6	55873889	51.063	45.635
Target Compounds							
2)	A alpha-BHC	4.013	3.278	188.9E6	94662123	49.345	52.899
3)	MA gamma-BHC...	4.347	3.608	169.0E6	85265927	48.098	51.405
4)	MA Heptachlor	4.939	3.951	159.8E6	85866180	46.439	53.370
5)	MB Aldrin	5.283	4.232	160.7E6	83190383	47.522	50.854
6)	B beta-BHC	4.540	3.904	66816812	36581952	44.461	49.755
7)	B delta-BHC	4.790	4.136	126.0E6	84301982	44.935	50.586
8)	B Heptachlo...	5.710	4.735	114.5E6	74073824	40.208	49.691
9)	A Endosulfan I	6.097	5.106	139.4E6	68519127	48.263	48.850
10)	B gamma-Chl...	5.967	4.986	145.1E6	74412238	47.389	48.831
11)	B alpha-Chl...	6.046	5.050	147.7E6	73395630	47.924	48.720
12)	B 4,4'-DDE	6.218	5.241	141.1E6	68129205	49.680	49.115
13)	MA Dieldrin	6.373	5.371	151.4E6	70938592	49.913	48.038
14)	MA Endrin	6.603	5.648	129.8E6	62665563	49.468	47.966
15)	B Endosulfa...	6.821	5.941	129.6E6	56114273	44.554	43.606
16)	A 4,4'-DDD	6.735	5.796	120.0E6	54687652	51.222	47.391
17)	MA 4,4'-DDT	7.051	6.048	138.3E6	58453613	50.704	48.803
18)	B Endrin al...	6.951	6.121	102.5E6	45192390	49.444	44.412
19)	B Endosulfa...	7.187	6.344	129.4E6	57198939	50.217	47.154
20)	A Methoxychlor	7.526	6.624	75793432	32694299	51.034	49.983
21)	B Endrin ke...	7.674	6.851	149.7E6	67188186	51.075	46.228
22)	Mirex	8.150	7.035	111.7E6	53527722	49.550	46.394

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD030323\
Data File : PD074110.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Mar 2023 11:50
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 02 21:29:07 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD030123.M
Quant Title : GC Extractables
QLast Update : Wed Mar 01 01:35:06 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

