

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD030323\
 Data File : PD074125.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Mar 2023 15:15
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 02 21:31:18 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD030123.M
 Quant Title : GC Extractables
 QLast Update : Wed Mar 01 01:35:06 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.552	2.774	125.0E6	63130759	54.550	53.301
2)	SA Decachlor...	9.100	7.930	138.2E6	59275303	51.848	48.413
Target Compounds							
2)	A alpha-BHC	4.009	3.278	221.5E6	98668021	57.857	55.138
3)	MA gamma-BHC...	4.343	3.608	200.0E6	90848921	56.919	54.771
4)	MA Heptachlor	4.936	3.951	191.0E6	88350095	55.496	54.914
5)	MB Aldrin	5.279	4.232	192.1E6	89189434	56.817	54.521
6)	B beta-BHC	4.537	3.904	82620301	39407978	54.977	53.599
7)	B delta-BHC	4.786	4.135	160.9E6	91884282	57.403	55.136
8)	B Heptachlo...	5.705	4.735	157.5E6	80541742	55.315	54.030
9)	A Endosulfan I	6.093	5.106	161.5E6	75567688	55.908	53.875
10)	B gamma-Chl...	5.963	4.986	172.1E6	82600402	56.205	54.204
11)	B alpha-Chl...	6.043	5.050	172.4E6	81222011	55.919	53.916
12)	B 4,4'-DDE	6.214	5.240	162.5E6	75984349	57.224	54.777
13)	MA Dieldrin	6.369	5.371	172.1E6	79999079	56.721	54.174
14)	MA Endrin	6.600	5.647	148.4E6	70409658	56.587	53.893
15)	B Endosulfa...	6.818	5.940	140.9E6	64171016	48.453	49.866
16)	A 4,4'-DDD	6.732	5.795	134.2E6	63335339	57.300	54.885
17)	MA 4,4'-DDT	7.048	6.047	143.5E6	63600142	52.619	53.100
18)	B Endrin al...	6.948	6.120	109.4E6	50478611	52.770	49.606
19)	B Endosulfa...	7.183	6.343	139.6E6	63945350	54.183	52.716
20)	A Methoxychlor	7.522	6.623	76638988	34924405	51.604	53.392
21)	B Endrin ke...	7.670	6.850	157.1E6	74573730	53.583	51.309
22)	Mirex	8.147	7.035	116.1E6	58019981	51.510	50.288

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD030323\
Data File : PD074125.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Mar 2023 15:15
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 02 21:31:18 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD030123.M
Quant Title : GC Extractables
QLast Update : Wed Mar 01 01:35:06 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

