

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD030421\
 Data File : PD061430.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Mar 2021 19:52
 Operator : DD\AJ
 Sample : INDB301
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 05 05:12:27 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD030421CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 05 05:11:33 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.487	3.982	17986494	184.4E6	20.000	20.000
27)	SA Decachlor...	8.261	9.041	53716234	385.1E6	40.000	40.000
Target Compounds							
5)	MB Aldrin	4.765	5.469	25783761	265.8E6	20.000	20.000
6)	B beta-BHC	4.462	4.825	12994515	114.9E6	20.000	20.000
7)	B delta-BHC	4.668	5.042	26452895	264.4E6	20.000	20.000
8)	B Heptachlo...	5.210	5.855	25468623	237.3E6	20.000	20.000
10)	B trans-Chl...	5.438	6.088	26551049	256.7E6	20.000	20.000
11)	B cis-Chlor...	5.497	6.162	25995875	243.5E6	20.000	20.000
12)	B 4,4'-DDE	5.660	6.318	50180960	491.3E6	40.000	40.000
15)	B Endosulfa...	6.326	6.894	44867111	396.3E6	40.000	40.000
18)	B Endrin al...	6.495	7.018	38683922	338.0E6	40.000	40.000
19)	B Endosulfa...	6.709	7.241	45522347	392.1E6	40.000	40.000
21)	B Endrin ke...	7.203	7.711	52723452	389.0E6	40.000	40.000


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD030421\
Data File : PD061430.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Mar 2021 19:52
Operator : DD\AJ
Sample : INDB301
Misc :
ALS Vial : 14 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 05 05:12:27 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD030421CLP.M
Quant Title : GC Extractables
QLast Update : Fri Mar 05 05:11:33 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 3.487 min
Delta R.T.: 0.000 min
Response: 17986494
Conc: 20.00 ng/ml

Instrument :
ECD_D
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.982 min
Delta R.T.: 0.000 min
Response: 184401037
Conc: 20.00 ng/ml




#5 Aldrin

R.T.: 4.765 min
Delta R.T.: 0.000 min
Response: 25783761
Conc: 20.00 ng/ml



#5 Aldrin


R.T.: 5.469 min
Delta R.T.: 0.000 min
Response: 265822631
Conc: 20.00 ng/ml



#6 beta-BHC


R.T.: 4.462 min
Delta R.T.: 0.000 min
Response: 12994515
Conc: 20.00 ng/ml

Instrument :
ECD_D
ClientSampleId :




#6 beta-BHC

R.T.: 4.825 min
Delta R.T.: 0.000 min
Response: 114933796
Conc: 20.00 ng/ml




#7 delta-BHC

R.T.: 4.668 min
Delta R.T.: 0.000 min
Response: 26452895
Conc: 20.00 ng/ml



#7 delta-BHC


R.T.: 5.042 min
Delta R.T.: 0.000 min
Response: 264428299
Conc: 20.00 ng/ml



#8 Heptachlor epoxide


R.T.: 5.210 min
Delta R.T.: 0.000 min
Response: 25468623
Conc: 20.00 ng/ml

Instrument :
ECD_D
ClientSampleId :




#8 Heptachlor epoxide

R.T.: 5.855 min
Delta R.T.: 0.000 min
Response: 237287139
Conc: 20.00 ng/ml




#10 trans-Chlordane

R.T.: 5.438 min
Delta R.T.: 0.000 min
Response: 26551049
Conc: 20.00 ng/ml



#10 trans-Chlordane


R.T.: 6.088 min
Delta R.T.: 0.000 min
Response: 256722246
Conc: 20.00 ng/ml



#11 cis-Chlordane


R.T.: 5.497 min
Delta R.T.: 0.000 min
Response: 25995875
Conc: 20.00 ng/ml

Instrument :
ECD_D
ClientSampleId :




#11 cis-Chlordane

R.T.: 6.162 min
Delta R.T.: 0.000 min
Response: 243493514
Conc: 20.00 ng/ml




#12 4,4'-DDE

R.T.: 5.660 min
Delta R.T.: 0.000 min
Response: 50180960
Conc: 40.00 ng/ml



#12 4,4'-DDE


R.T.: 6.318 min
Delta R.T.: 0.000 min
Response: 491319062
Conc: 40.00 ng/ml



#15 Endosulfan II


R.T.: 6.326 min
Delta R.T.: 0.000 min
Response: 44867111
Conc: 40.00 ng/ml

Instrument :
ECD_D
ClientSampleId :




#15 Endosulfan II

R.T.: 6.894 min
Delta R.T.: 0.000 min
Response: 396282899
Conc: 40.00 ng/ml




#18 Endrin aldehyde

R.T.: 6.495 min
Delta R.T.: 0.000 min
Response: 38683922
Conc: 40.00 ng/ml



#18 Endrin aldehyde


R.T.: 7.018 min
Delta R.T.: 0.000 min
Response: 337977196
Conc: 40.00 ng/ml



#19 Endosulfan Sulfate


R.T.: 6.709 min
Delta R.T.: 0.000 min
Response: 45522347
Conc: 40.00 ng/ml

Instrument :
ECD_D
ClientSampleId :




#19 Endosulfan Sulfate

R.T.: 7.241 min
Delta R.T.: 0.000 min
Response: 392077125
Conc: 40.00 ng/ml




#21 Endrin ketone

R.T.: 7.203 min
Delta R.T.: 0.000 min
Response: 52723452
Conc: 40.00 ng/ml



#21 Endrin ketone


R.T.: 7.711 min
Delta R.T.: 0.000 min
Response: 389034999
Conc: 40.00 ng/ml



#27 Decachlorobiphenyl

R.T.: 8.261 min
Delta R.T.: 0.000 min
Response: 53716234
Conc: 40.00 ng/ml

Instrument :
ECD_D
ClientSampleId :



#27 Decachlorobiphenyl

R.T.: 9.041 min
Delta R.T.: 0.000 min
Response: 385145458
Conc: 40.00 ng/ml