

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD030422\
 Data File : PD068306.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Mar 2022 13:25
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 04 01:06:45 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD030122.M
 Quant Title : GC Extractables
 QLast Update : Tue Mar 01 06:48:49 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.450	4.176	59828349	483.7E6	51.037	50.394
28)	SA Decachlor...	8.214	9.373	62016557	315.9E6	53.292	53.276
Target Compounds							
2)	A alpha-BHC	3.889	4.572	91823590	712.2E6	53.513	51.907
3)	MA gamma-BHC...	4.173	4.863	84530797	644.2E6	53.082	51.954
4)	MA Heptachlor	4.472	5.390	80172901	595.6E6	53.107	52.585
5)	MB Aldrin	4.720	5.704	82955556	611.8E6	52.966	51.745
6)	B beta-BHC	4.423	5.025	36174494	275.3E6	51.556	50.929
7)	B delta-BHC	4.628	5.251	85614528	643.5E6	53.062	51.740
8)	B Heptachlo...	5.165	6.091	76497501	546.1E6	52.472	50.956
9)	A Endosulfan I	5.504	6.455	71207815	499.4E6	52.214	51.759
10)	B gamma-Chl...	5.391	6.328	77528104	548.9E6	52.079	51.915
11)	B alpha-Chl...	5.450	6.403	76592630	532.8E6	52.125	51.689
12)	B 4,4'-DDE	5.616	6.553	72422290	507.7E6	53.095	51.932
13)	MA Dieldrin	5.748	6.713	77462109	533.4E6	52.678	51.762
14)	MA Endrin	6.006	6.933	66330797	401.9E6	53.695	52.259
15)	B Endosulfa...	6.280	7.137	66993170	427.0E6	52.654	51.965
16)	A 4,4'-DDD	6.134	7.044	62329858	408.1E6	52.558	51.860
17)	MA 4,4'-DDT	6.375	7.347	61307373	385.6E6	55.594	54.495
18)	B Endrin al...	6.450	7.262	46990377	307.7E6	51.271	51.136
19)	B Endosulfa...	6.662	7.486	65790165	402.2E6	52.206	51.692
20)	A Methoxychlor	6.926	7.800	35119857	192.2E6	54.122	55.326
21)	B Endrin ke...	7.156	7.962	77096225	432.6E6	52.366	51.528
22)	Mirex	7.341	8.428	64086118	326.9E6	50.561	51.105

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD030422\
Data File : PD068306.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Mar 2022 13:25
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 04 01:06:45 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD030122.M
Quant Title : GC Extractables
QLast Update : Tue Mar 01 06:48:49 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

