

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD030424\
Data File : PD081917.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Mar 2024 17:09
Operator : AR\AJ
Sample : P1605-05
Misc :
ALS Vial : 21 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
BGRJ2

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 03/06/2024
Supervised By :Ankita Jodhani 03/06/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 04 21:18:37 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD021224CLP.M
Quant Title : GC Extractables
QLast Update : Mon Feb 12 17:06:05 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(6) beta-BHC (B)

0.000min 0.000 ng/ml

response 0

(6) beta-BHC #2 (B)

3.940min 0.333 ng/ml

response 562243

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD030424\
Data File : PD081917.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Mar 2024 17:09
Operator : AR\AJ
Sample : P1605-05
Misc :
ALS Vial : 21 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
BGRJ2

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 03/06/2024
Supervised By :Ankita Jodhani 03/06/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 04 21:18:37 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD021224CLP.M
Quant Title : GC Extractables
QLast Update : Mon Feb 12 17:06:05 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(6) beta-BHC (B)

4.557min 0.500 ng/ml m

response 519825

(6) beta-BHC #2 (B)

3.940min 0.333 ng/ml

response 562243

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD030424\
Data File : PD081917.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Mar 2024 17:09
Operator : AR\AJ
Sample : P1605-05
Misc :
ALS Vial : 21 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
BGRJ2

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 03/06/2024
Supervised By :Ankita Jodhani 03/06/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 04 21:18:37 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD021224CLP.M
Quant Title : GC Extractables
QLast Update : Mon Feb 12 17:06:05 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(13) Dieldrin (MA)

0.000min 0.000 ng/ml

response 0

(13) Dieldrin #2 (MA)

5.374min 0.782 ng/ml

response 2567173

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD030424\
Data File : PD081917.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Mar 2024 17:09
Operator : AR\AJ
Sample : P1605-05
Misc :
ALS Vial : 21 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
BGRJ2

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 03/06/2024
Supervised By :Ankita Jodhani 03/06/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 04 21:18:37 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD021224CLP.M
Quant Title : GC Extractables
QLast Update : Mon Feb 12 17:06:05 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(13) Dieldrin (MA)

6.376min 0.418 ng/ml m

response 929004

(13) Dieldrin #2 (MA)

5.372min 0.745 ng/ml m

response 2445873

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD030424\
Data File : PD081917.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Mar 2024 17:09
Operator : AR\AJ
Sample : P1605-05
Misc :
ALS Vial : 21 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
BGRJ2

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 03/06/2024
Supervised By :Ankita Jodhani 03/06/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 04 21:18:37 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD021224CLP.M
Quant Title : GC Extractables
QLast Update : Mon Feb 12 17:06:05 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(19) Endosulfan Sulfate (B)

0.000min 0.000 ng/ml

response 0

(19) Endosulfan Sulfate #2 (B)

6.380min 0.516 ng/ml

response 1385446

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD030424\
Data File : PD081917.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Mar 2024 17:09
Operator : AR\AJ
Sample : P1605-05
Misc :
ALS Vial : 21 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
BGRJ2

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 03/06/2024
Supervised By :Ankita Jodhani 03/06/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 04 21:18:37 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD021224CLP.M
Quant Title : GC Extractables
QLast Update : Mon Feb 12 17:06:05 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(19) Endosulfan Sulfate (B)

7.185min 0.599 ng/ml m

response 1096108

(19) Endosulfan Sulfate #2 (B)

6.378min 0.481 ng/ml m

response 1291023

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD030424\
Data File : PD081917.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Mar 2024 17:09
Operator : AR\AJ
Sample : P1605-05
Misc :
ALS Vial : 21 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
BGRJ2

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 03/06/2024
Supervised By :Ankita Jodhani 03/06/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 04 21:18:37 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD021224CLP.M
Quant Title : GC Extractables
QLast Update : Mon Feb 12 17:06:05 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(27) Decachlorobiphenyl (SA)

9.108min 14.973 ng/ml

response 28887824

(27) Decachlorobiphenyl #2 (SA)

7.924min 16.014 ng/ml

response 39726000

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD030424\
Data File : PD081917.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Mar 2024 17:09
Operator : AR\AJ
Sample : P1605-05
Misc :
ALS Vial : 21 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
BGRJ2

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 03/06/2024
Supervised By :Ankita Jodhani 03/06/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 04 21:18:37 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD021224CLP.M
Quant Title : GC Extractables
QLast Update : Mon Feb 12 17:06:05 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



QEdit

(27) Decachlorobiphenyl (SA)

9.108min 14.973 ng/ml

response 28887824

(27) Decachlorobiphenyl #2 (SA)

7.923min 15.656 ng/ml m

response 38836742