

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD030424\
 Data File : PD081933.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Mar 2024 20:42
 Operator : AR\AJ
 Sample : P1605-18
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 BGSM7

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 03/05/2024
 Supervised By : Ankita Jodhani 03/05/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 04 21:22:06 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD021224CLP.M
 Quant Title : GC Extractables
 QLast Update : Mon Feb 12 17:06:05 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.562	2.789	20790817	35510854	15.208	14.485m
2)	SA Decachlor...	9.105	7.923	27982592	37736240	14.504	15.212
Target Compounds							
9)	A Endosulfan I	6.083	5.101	5620721	2033715	2.669	0.679 #
13)	MA Dieldrin	6.375	5.374	18805498	23670723	8.469	7.207m
19)	B Endosulfa...	7.185	6.377	1832418	1393123	1.001	0.519m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD030424\
Data File : PD081933.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Mar 2024 20:42
Operator : AR\AJ
Sample : P1605-18
Misc :
ALS Vial : 34 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 04 21:22:06 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD021224CLP.M
Quant Title : GC Extractables
QLast Update : Mon Feb 12 17:06:05 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Instrument :
ECD_D
Client Sampled :
BGSMT7

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 03/05/2024
Supervised By : Ankita Jodhani 03/05/2024