

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD030519\
Data File : PD051612.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 05 Mar 2019 15:30
Operator : AJ\SJ
Sample : K1626-01
Misc :
ALS Vial : 19 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
BF663

Manual Integrations
APPROVED

Sohil
3/6/2019 9:54:15 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 06 04:15:57 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD022219CLP.M
Quant Title : GC Extractables
QLast Update : Fri Feb 22 04:35:08 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



QEdit

(13) Dieldrin (MA)
5.407min 1.756 ng/ml
response 4065234

(13) Dieldrin #2 (MA)
6.549min 1.606 ng/ml
response 53450077

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD030519\
Data File : PD051612.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 05 Mar 2019 15:30
Operator : AJ\SJ
Sample : K1626-01
Misc :
ALS Vial : 19 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
BF663

Manual Integrations
APPROVED

Sohil
3/6/2019 9:54:15 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 06 04:15:57 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD022219CLP.M
Quant Title : GC Extractables
QLast Update : Fri Feb 22 04:35:08 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



QEdit

(13) Dieldrin (MA)
5.406min 2.050 ng/ml m
response 4745637

(13) Dieldrin #2 (MA)
6.549min 1.606 ng/ml
response 53450077

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD030519\
Data File : PD051612.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 05 Mar 2019 15:30
Operator : AJ\SJ
Sample : K1626-01
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
BF663

Manual Integrations
APPROVED

Sohil
3/6/2019 9:54:15 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 06 04:15:57 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD022219CLP.M
Quant Title : GC Extractables
QLast Update : Fri Feb 22 04:35:08 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.224	4.044	15103022	350.5E6	10.510	11.344
27) SA Decachlor...	7.825	9.159	18850159	295.8E6	15.674	15.603
Target Compounds						
13) MA Dieldrin	5.406	6.549	4745637	53450077	2.050m)	1.606

2550311/19

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.