

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD030620\
 Data File : PD057519.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Mar 2020 21:49
 Operator : AJ\MA
 Sample : PEM41
 Misc :
 ALS Vial : 100 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PEM41

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 09 07:56:56 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD030620CLP.M
 Quant Title : GC Extractables
 QLast Update : Mon Mar 09 05:47:43 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.228	3.898	19462884	44353150	18.394	18.423
27)	SA Decachlor...	7.869	8.900	23696686	49808163	19.252	19.233
Target Compounds							
2)	A alpha-BHC	3.651	4.281	13690479	32651101	8.739	9.136
3)	MA gamma-BHC...	3.923	4.562	13807217	31443257	8.877	8.919
6)	B beta-BHC	4.171	4.732	6321929	15519231	9.523	9.317
12)	B 4,4'-DDE	5.317	6.210	223900	587848	0.167	0.173
14)	MA Endrin	5.688	6.571	56762515	136.1E6	46.488	46.344
16)	A 4,4'-DDD	5.825	6.697	1295779	3414344	1.150	1.215
17)	MA 4,4'-DDT	6.059	6.994	112.7E6	264.6E6	96.342	94.056
18)	B Endrin al...	6.130	6.904	568295	1487653	0.555	0.589
20)	A Methoxychlor	6.607	7.452	136.7E6	330.6E6	232.160	227.650
21)	B Endrin ke...	6.819	7.592	2734826	4233847	1.942	1.396 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD030620\
Data File : PD057519.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Mar 2020 21:49
Operator : AJ\MA
Sample : PEM41
Misc :
ALS Vial : 100 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 09 07:56:56 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD030620CLP.M
Quant Title : GC Extractables
QLast Update : Mon Mar 09 05:47:43 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

