

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ030822.M
 Title : GC EXTRACTABLES
 Last Update : Wed Mar 09 00:56:32 2022
 Response Via : Initial Calibration

Calibration Files

500 =PQ056348.D 1000 =PQ056346.D 750 =PQ056347.D
 250 =PQ056349.D 50 =PQ056350.D

Compound			500	1000	750	250	50	Avg	%RSD	
1)	SA	Tetrachloro-m-xylene	1.227	1.177	1.203	1.254	1.330	1.238	E7	4.76
2)	SA	Decachlorobiphenyl	1.490	1.382	1.433	1.555	1.673	1.507	E7	7.52
3)	L1	AR-1016-1	3.058	2.918	2.940	3.247	3.510	3.135	E5	7.88
4)	L1	AR-1016-2	6.194	5.678	5.974	6.431	7.694	6.394	E5	12.17
5)	L1	AR-1016-3	3.999	3.614	3.768	4.177	4.787	4.069	E5	11.19
6)	L1	AR-1016-4	3.291	3.014	3.135	3.452	3.949	3.368	E5	10.81
7)	L1	AR-1016-5	2.633	2.414	2.522	2.731	3.332	2.726	E5	13.16
8)	L2	AR-1221-1						0.000		-1.00
9)	L2	AR-1221-2						0.000		-1.00
10)	L2	AR-1221-3						0.000		-1.00
11)	L3	AR-1232-1						0.000		-1.00
12)	L3	AR-1232-2						0.000		-1.00
13)	L3	AR-1232-3						0.000		-1.00
14)	L3	AR-1232-4						0.000		-1.00
15)	L3	AR-1232-5						0.000		-1.00
16)	L4	AR-1242-1						0.000		-1.00
17)	L4	AR-1242-2						0.000		-1.00
18)	L4	AR-1242-3						0.000		-1.00
19)	L4	AR-1242-4						0.000		-1.00
20)	L4	AR-1242-5						0.000		-1.00
21)	L5	AR-1248-1						0.000		-1.00
22)	L5	AR-1248-2						0.000		-1.00
23)	L5	AR-1248-3						0.000		-1.00
24)	L5	AR-1248-4						0.000		-1.00
25)	L5	AR-1248-5						0.000		-1.00
26)	L6	AR-1254-1						0.000		-1.00
27)	L6	AR-1254-2						0.000		-1.00
28)	L6	AR-1254-3						0.000		-1.00
29)	L6	AR-1254-4						0.000		-1.00
30)	L6	AR-1254-5						0.000		-1.00
31)	L7	AR-1260-1	4.477	4.142	4.296	4.742	4.938	4.519	E5	7.16
32)	L7	AR-1260-2	5.643	5.204	5.418	5.992	6.750	5.802	E5	10.43
33)	L7	AR-1260-3	4.686	4.231	4.416	5.080	4.731	4.629	E5	7.01
34)	L7	AR-1260-4	5.490	5.009	5.153	5.904	5.682	5.448	E5	6.77
35)	L7	AR-1260-5	1.253	1.190	1.215	1.325	1.273	1.251	E6	4.19
36)	L8	AR-1262-1						0.000		-1.00
37)	L8	AR-1262-2						0.000		-1.00
38)	L8	AR-1262-3						0.000		-1.00
39)	L8	AR-1262-4						0.000		-1.00
40)	L8	AR-1262-5						0.000		-1.00
41)	L9	AR-1268-1						0.000		-1.00
42)	L9	AR-1268-2						0.000		-1.00
43)	L9	AR-1268-3						0.000		-1.00
44)	L9	AR-1268-4						0.000		-1.00
45)	L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

500 =PQ056348.D 1000 =PQ056346.D 750 =PQ056347.D
 250 =PQ056349.D 50 =PQ056350.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	1.204	1.185	1.201	1.180	1.128	1.180	E7	2.61
2) SA	Decachlorobiphenyl	8.947	8.640	8.815	9.068	9.478	8.990	E6	3.52
3) L1	AR-1016-1	3.340	3.272	3.348	3.231	3.197	3.278	E5	2.01
4) L1	AR-1016-2	6.332	6.104	6.207	6.508	6.602	6.351	E5	3.24
5) L1	AR-1016-3	2.776	2.689	2.738	2.804	2.821	2.766	E5	1.93

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ030822.M
 Title : GC EXTRACTABLES
 Last Update : Wed Mar 09 00:56:32 2022
 Response Via : Initial Calibration

Calibration Files

500 =PQ056348.D 1000 =PQ056346.D 750 =PQ056347.D
 250 =PQ056349.D 50 =PQ056350.D

	Compound	500	1000	750	250	50	Avg		%RSD

6)	L1 AR-1016-4	2.425	2.249	2.330	2.535	2.665	2.441	E5	6.74
7)	L1 AR-1016-5	2.881	2.755	2.826	2.949	3.060	2.894	E5	4.04
8)	L2 AR-1221-1						0.000		-1.00
9)	L2 AR-1221-2						0.000		-1.00
10)	L2 AR-1221-3						0.000		-1.00
11)	L3 AR-1232-1						0.000		-1.00
12)	L3 AR-1232-2						0.000		-1.00
13)	L3 AR-1232-3						0.000		-1.00
14)	L3 AR-1232-4						0.000		-1.00
15)	L3 AR-1232-5						0.000		-1.00
16)	L4 AR-1242-1						0.000		-1.00
17)	L4 AR-1242-2						0.000		-1.00
18)	L4 AR-1242-3						0.000		-1.00
19)	L4 AR-1242-4						0.000		-1.00
20)	L4 AR-1242-5						0.000		-1.00
21)	L5 AR-1248-1						0.000		-1.00
22)	L5 AR-1248-2						0.000		-1.00
23)	L5 AR-1248-3						0.000		-1.00
24)	L5 AR-1248-4						0.000		-1.00
25)	L5 AR-1248-5						0.000		-1.00
26)	L6 AR-1254-1						0.000		-1.00
27)	L6 AR-1254-2						0.000		-1.00
28)	L6 AR-1254-3						0.000		-1.00
29)	L6 AR-1254-4						0.000		-1.00
30)	L6 AR-1254-5						0.000		-1.00
31)	L7 AR-1260-1	5.767	5.505	5.667	5.898	6.641	5.896	E5	7.48
32)	L7 AR-1260-2	6.819	6.561	6.724	6.936	7.331	6.874	E5	4.22
33)	L7 AR-1260-3	6.375	6.197	6.324	6.424	6.341	6.332	E5	1.34
34)	L7 AR-1260-4	4.750	4.594	4.700	4.823	5.263	4.826	E5	5.35
35)	L7 AR-1260-5	1.104	1.088	1.102	1.099	1.078	1.094	E6	0.99
36)	L8 AR-1262-1						0.000		-1.00
37)	L8 AR-1262-2						0.000		-1.00
38)	L8 AR-1262-3						0.000		-1.00
39)	L8 AR-1262-4						0.000		-1.00
40)	L8 AR-1262-5						0.000		-1.00
41)	L9 AR-1268-1						0.000		-1.00
42)	L9 AR-1268-2						0.000		-1.00
43)	L9 AR-1268-3						0.000		-1.00
44)	L9 AR-1268-4						0.000		-1.00
45)	L9 AR-1268-5						0.000		-1.00

(#) = Out of Range