

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD030822\
 Data File : PD068384.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Mar 2022 10:44
 Operator : AR\AJ
 Sample : N1806-07DL 2X
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 DUM-25DL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 08 23:51:06 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD030122.M
 Quant Title : GC Extractables
 QLast Update : Tue Mar 01 06:48:49 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.450	4.175	13840782	114.6E6	11.807	11.941
28)	SA Decachlor...	8.213	9.371	13426130	63052485	11.537	10.635
Target Compounds							
10)	B gamma-Chl...	5.391	6.326	98967094	724.3E6	66.481	68.498
11)	B alpha-Chl...	5.450	6.402	96816321	811.0E6	65.888	78.677



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD030822\
Data File : PD068384.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Mar 2022 10:44
Operator : AR\AJ
Sample : N1806-07DL 2X
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
DUM-25DL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 08 23:51:06 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD030122.M
Quant Title : GC Extractables
QLast Update : Tue Mar 01 06:48:49 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 3.450 min
Delta R.T.: 0.000 min
Response: 13840782
Conc: 11.81 ng/ml

Instrument :
ECD_D
ClientSampleId :
DUM-25DL




#1 Tetrachloro-m-xylene

R.T.: 4.175 min
Delta R.T.: -0.001 min
Response: 114612470
Conc: 11.94 ng/ml




#10 gamma-Chlordane

R.T.: 5.391 min
Delta R.T.: 0.000 min
Response: 98967094
Conc: 66.48 ng/ml



#10 gamma-Chlordane


R.T.: 6.326 min
Delta R.T.: -0.002 min
Response: 724291882
Conc: 68.50 ng/ml



#11 alpha-Chlordane


R.T.: 5.450 min
Delta R.T.: 0.000 min
Response: 96816321
Conc: 65.89 ng/ml

Instrument :
ECD_D
ClientSampleId :
DUM-25DL




#11 alpha-Chlordane

R.T.: 6.402 min
Delta R.T.: 0.000 min
Response: 810955826
Conc: 78.68 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.213 min
Delta R.T.: 0.000 min
Response: 13426130
Conc: 11.54 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.371 min
Delta R.T.: -0.002 min
Response: 63052485
Conc: 10.63 ng/ml