

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD030922\
 Data File : PD068426.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Mar 2022 16:59
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 10 01:25:57 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD030122.M
 Quant Title : GC Extractables
 QLast Update : Tue Mar 01 06:48:49 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.449	4.176	62695442	480.7E6	53.483	50.080
28)	SA Decachlor...	8.211	9.373	62009749	296.1E6	53.286	49.940
Target Compounds							
2)	A alpha-BHC	3.889	4.572	96599262	711.2E6	56.296	51.828
3)	MA gamma-BHC...	4.172	4.862	88893989	639.2E6	55.822	51.547
4)	MA Heptachlor	4.471	5.390	83428245	586.8E6	55.264	51.807
5)	MB Aldrin	4.719	5.703	86653547	601.0E6	55.328	50.834
6)	B beta-BHC	4.422	5.025	37671522	273.5E6	53.689	50.601
7)	B delta-BHC	4.627	5.251	90378484	640.3E6	56.014	51.483
8)	B Heptachlo...	5.164	6.091	79990055	546.5E6	54.867	50.994
9)	A Endosulfan I	5.503	6.454	74785663	489.5E6	54.837	50.731
10)	B gamma-Chl...	5.390	6.328	82136906	541.4E6	55.175	51.197
11)	B alpha-Chl...	5.449	6.403	80592651	524.5E6	54.847	50.888
12)	B 4,4'-DDE	5.615	6.552	76938874	497.1E6	56.406	50.845
13)	MA Dieldrin	5.747	6.712	81486193	521.8E6	55.415	50.639
14)	MA Endrin	6.005	6.932	63706395	352.6E6	51.570	45.851
15)	B Endosulfa...	6.279	7.138	70451245	416.6E6	55.372	50.698
16)	A 4,4'-DDD	6.133	7.044	67460421	411.3E6	56.884	52.260
17)	MA 4,4'-DDT	6.373	7.346	62200042	368.6E6	56.403	52.093
18)	B Endrin al...	6.449	7.262	51244257	311.1E6	55.912	51.702
19)	B Endosulfa...	6.661	7.486	69902205	396.2E6	55.469	50.923
20)	A Methoxychlor	6.925	7.800	34712676	179.9E6	53.495	51.768
21)	B Endrin ke...	7.155	7.962	84143708	434.9E6	57.152	51.810
22)	Mirex	7.339	8.428	66966832	313.9E6	52.833	49.075

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD030922\
Data File : PD068426.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Mar 2022 16:59
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 10 01:25:57 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD030122.M
Quant Title : GC Extractables
QLast Update : Tue Mar 01 06:48:49 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

