

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD030922\
 Data File : PD068439.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Mar 2022 22:27
 Operator : AR\AJ
 Sample : N1830-02MS
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 11986MS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 10 02:08:22 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD030122.M
 Quant Title : GC Extractables
 QLast Update : Tue Mar 01 06:48:49 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.450	4.176	25815161	207.8E6	22.022	21.653
28)	SA Decachlor...	8.212	9.373	22132036	106.5E6	19.019	17.965
Target Compounds							
2)	A alpha-BHC	3.889	4.572	114.8E6	805.6E6	66.919	58.709
3)	MA gamma-BHC...	4.173	4.863	106.0E6	738.9E6	66.583	59.593
4)	MA Heptachlor	4.471	5.390	104.1E6	707.8E6	68.955	62.490
5)	MB Aldrin	4.720	5.703	101.8E6	681.2E6	64.990	57.612
6)	B beta-BHC	4.423	5.025	42665927	308.1E6	60.807	56.991
7)	B delta-BHC	4.627	5.251	106.3E6	721.8E6	65.882	58.032
8)	B Heptachlo...	5.164	6.091	94957597	618.7E6	65.134	57.729
9)	A Endosulfan I	5.502	6.454	88855568	568.7E6	65.154	58.942
10)	B gamma-Chl...	5.390	6.327	96512129	618.6E6	64.831	58.502
11)	B alpha-Chl...	5.449	6.402	94774086	600.2E6	64.498	58.229
12)	B 4,4'-DDE	5.615	6.552	89749611	573.0E6	65.798	58.616
13)	MA Dieldrin	5.747	6.712	94897996	596.5E6	64.535	57.882
14)	MA Endrin	6.004	6.932	85920437	489.6E6	69.552	63.662
15)	B Endosulfa...	6.279	7.137	78619449	466.7E6	61.792	56.791
16)	A 4,4'-DDD	6.133	7.043	76045761	461.0E6	64.123	58.577
17)	MA 4,4'-DDT	6.373	7.346	79472220	475.7E6	72.066	67.226
18)	B Endrin al...	6.448	7.261	60903329	370.7E6	66.451	61.601
19)	B Endosulfa...	6.662	7.486	80844712	459.5E6	64.152	59.060
20)	A Methoxychlor	6.924	7.799	43669824	230.1E6	67.298	66.224
21)	B Endrin ke...	7.154	7.961	91697707	482.0E6	62.283	57.420
22)	Mirex	7.340	8.427	79333441	379.9E6	62.590	59.400

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD030922\
Data File : PD068439.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Mar 2022 22:27
Operator : AR\AJ
Sample : N1830-02MS
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
11986MS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 10 02:08:22 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD030122.M
Quant Title : GC Extractables
QLast Update : Tue Mar 01 06:48:49 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

