

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : DTX031021.M
 Title : GC Extractables
 Last Update : Thu Mar 11 03:56:20 2021
 Response Via : Initial Calibration

Calibration Files

500 =PD061517.D 1000 =PD061515.D 750 =PD061516.D
 250 =PD061518.D 100 =PD061519.D

	Compound	500	1000	750	250	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	8.302	8.170	8.083	8.094	9.485	8.427	E5	7.10
2)	Toxaphene-1	1.236	1.175	1.215	1.138	1.435	1.240	E4	9.32
3)	Toxaphene-2	1.022	0.932	0.960	1.007	1.105	1.005	E4	6.61
4)	Toxaphene-3	2.169	2.098	2.076	2.136	2.189	2.133	E4	2.21
5)	Toxaphene-4	8.128	7.736	7.958	7.643	8.365	7.966	E3	3.67
6)	Toxaphene-5	0.909	0.842	0.857	0.886	1.238	0.947	E4	17.43
7) SA	Decachlorobiphenyl	1.202	1.142	1.160	1.270	1.536	1.262	E6	12.73

Signal #2 Calibration Files

500 =PD061517.D 1000 =PD061515.D 750 =PD061516.D
 250 =PD061518.D 100 =PD061519.D

	Compound	500	1000	750	250	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	0.868	0.821	0.839	0.843	1.005	0.875	E7	8.51
2)	Toxaphene-1	1.078	1.073	1.099	0.956	1.193	1.080	E5	7.81
3)	Toxaphene-2	9.533	9.240	9.032	8.849	9.024	9.136	E4	2.87
4)	Toxaphene-3	2.643	2.540	2.550	2.504	3.070	2.661	E5	8.80
5)	Toxaphene-4	1.643	1.599	1.616	1.682	2.084	1.725	E5	11.77
6)	Toxaphene-5	1.141	1.097	1.107	1.101	1.308	1.151	E5	7.79
7) SA	Decachlorobiphenyl	0.870	0.812	0.832	0.901	1.076	0.898	E7	11.72

(#) = Out of Range