

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD031021\
 Data File : PD061508.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Mar 2021 11:28
 Operator : AR\AJ
 Sample : PSTDICC025
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDICC025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 10 12:31:43 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031021.M
 Quant Title : GC Extractables
 QLast Update : Wed Mar 10 12:30:41 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.487	3.982	20882640	214.4E6	24.826	24.490
28)	SA Decachlor...	8.257	9.037	32394171	231.7E6	26.442	26.313
Target Compounds							
2)	A alpha-BHC	3.926	4.370	30827944	340.1E6	23.572	24.613
3)	MA gamma-BHC...	4.211	4.655	29768364	315.3E6	24.040	24.827
4)	MA Heptachlor	4.514	5.163	31915799	332.3E6	24.482	25.100
5)	MB Aldrin	4.764	5.468	30041559	312.6E6	24.276	24.964
6)	B beta-BHC	4.461	4.824	14258293	129.1E6	25.554	24.622
7)	B delta-BHC	4.667	5.041	30485453	307.4E6	23.673	24.213
8)	B Heptachlo...	5.208	5.853	29510142	277.7E6	24.840	24.672
9)	A Endosulfan I	5.549	6.209	28222438	270.7E6	24.888	25.170
10)	B gamma-Chl...	5.436	6.087	30030763	289.0E6	24.669	24.596
11)	B alpha-Chl...	5.494	6.159	30062858	282.3E6	24.871	24.971
12)	B 4,4'-DDE	5.657	6.316	28127311	287.0E6	24.168	24.948
13)	MA Dieldrin	5.793	6.466	29552211	287.1E6	24.440	25.021
14)	MA Endrin	6.051	6.684	27201739	243.8E6	24.716	24.885
15)	B Endosulfa...	6.324	6.892	26167720	235.9E6	24.958	25.492
16)	A 4,4'-DDD	6.174	6.805	24064420	234.3E6	24.156	25.059
17)	MA 4,4'-DDT	6.417	7.103	25974515	238.9E6	24.371	24.521
18)	B Endrin al...	6.493	7.016	21806812	189.2E6	25.549	25.120
19)	B Endosulfa...	6.707	7.238	27805117	236.4E6	25.377	24.520
20)	A Methoxychlor	6.962	7.560	15314206	117.3E6	25.344	24.629
21)	B Endrin ke...	7.200	7.708	32377591	242.9E6	25.021	25.454
22)	Mirex	7.391	8.163	26416537	163.7E6	26.634	26.267

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD031021\
 Data File : PD061508.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Mar 2021 11:28
 Operator : AR\AJ
 Sample : PSTDICC025
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_D
 Client Sampled :
 PSTDICC025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 10 12:31:43 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031021.M
 Quant Title : GC Extractables
 QLast Update : Wed Mar 10 12:30:41 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

