

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD031121\
 Data File : PD061526.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Mar 2021 09:31
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 12 07:32:30 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031021.M
 Quant Title : GC Extractables
 QLast Update : Thu Mar 11 04:34:41 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.490	3.982	44907975	462.5E6	56.025	55.439
28)	SA Decachlor...	8.262	9.038	63831176	447.5E6	52.930	51.862
Target Compounds							
2)	A alpha-BHC	3.929	4.370	70087447	734.3E6	56.874	56.655
3)	MA gamma-BHC...	4.214	4.655	67388229	670.2E6	57.252	55.239
4)	MA Heptachlor	4.518	5.163	68722157	690.4E6	55.393	54.196
5)	MB Aldrin	4.768	5.468	66516740	664.3E6	56.786	55.549
6)	B beta-BHC	4.464	4.824	29520561	277.7E6	55.832	56.939
7)	B delta-BHC	4.671	5.041	68289765	674.6E6	57.790	56.254
8)	B Heptachlo...	5.212	5.853	63344170	611.3E6	55.730	56.168
9)	A Endosulfan I	5.553	6.210	60384188	568.1E6	55.582	54.608
10)	B gamma-Chl...	5.440	6.087	65005559	628.6E6	55.858	55.584
11)	B alpha-Chl...	5.498	6.160	64228968	594.8E6	55.487	54.557
12)	B 4,4'-DDE	5.661	6.316	62086186	605.3E6	56.538	54.502
13)	MA Dieldrin	5.797	6.467	64416554	603.7E6	56.130	54.631
14)	MA Endrin	6.055	6.684	54035221	475.0E6	51.164	50.666
15)	B Endosulfa...	6.328	6.893	55841388	483.8E6	55.875	54.109
16)	A 4,4'-DDD	6.179	6.806	54603982	497.4E6	58.767	55.553
17)	MA 4,4'-DDT	6.421	7.104	54922316	498.6E6	55.235	53.909
18)	B Endrin al...	6.497	7.017	45751022	397.9E6	55.395	55.398
19)	B Endosulfa...	6.710	7.239	57225104	498.3E6	55.045	54.095
20)	A Methoxychlor	6.967	7.562	31598407	239.2E6	56.057	51.702
21)	B Endrin ke...	7.204	7.710	71410760	504.9E6	57.797	54.731
22)	Mirex	7.395	8.165	51160996	320.3E6	51.602	52.733

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD031121\
Data File : PD061526.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Mar 2021 09:31
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 12 07:32:30 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031021.M
Quant Title : GC Extractables
QLast Update : Thu Mar 11 04:34:41 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

