

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD031122\
 Data File : PD068483.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Mar 2022 09:55
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 11 23:10:38 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD030122.M
 Quant Title : GC Extractables
 QLast Update : Tue Mar 01 06:48:49 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.450	4.175	61522309	474.5E6	52.482	49.437
28)	SA Decachlor...	8.211	9.370	63486553	292.4E6	54.555	49.322
Target Compounds							
2)	A alpha-BHC	3.889	4.571	94903905	703.2E6	55.308	51.249
3)	MA gamma-BHC...	4.173	4.862	87244777	634.4E6	54.786	51.162
4)	MA Heptachlor	4.472	5.389	83948510	595.9E6	55.608	52.612
5)	MB Aldrin	4.719	5.702	86216345	600.4E6	55.048	50.778
6)	B beta-BHC	4.423	5.024	36977571	271.1E6	52.700	50.147
7)	B delta-BHC	4.627	5.250	89123473	636.4E6	55.237	51.164
8)	B Heptachlo...	5.164	6.090	79092798	536.0E6	54.252	50.011
9)	A Endosulfan I	5.503	6.453	74181909	484.1E6	54.395	50.168
10)	B gamma-Chl...	5.390	6.326	81270964	535.4E6	54.593	50.633
11)	B alpha-Chl...	5.449	6.402	80247074	520.0E6	54.612	50.452
12)	B 4,4'-DDE	5.615	6.552	75637156	493.2E6	55.452	50.444
13)	MA Dieldrin	5.747	6.711	80607711	515.9E6	54.817	50.058
14)	MA Endrin	6.004	6.931	67572339	390.4E6	54.700	50.765
15)	B Endosulfa...	6.278	7.136	69434553	411.4E6	54.573	50.066
16)	A 4,4'-DDD	6.132	7.042	65121503	399.5E6	54.912	50.756
17)	MA 4,4'-DDT	6.373	7.345	64908199	381.9E6	58.859	53.973
18)	B Endrin al...	6.449	7.260	50315402	305.7E6	54.899	50.805
19)	B Endosulfa...	6.661	7.484	69220134	394.4E6	54.928	50.687
20)	A Methoxychlor	6.923	7.798	36684827	189.0E6	56.534	54.393
21)	B Endrin ke...	7.154	7.960	81420369	419.3E6	55.303	49.945
22)	Mirex	7.339	8.426	66843289	314.2E6	52.736	49.132

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD031122\
Data File : PD068483.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Mar 2022 09:55
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 11 23:10:38 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD030122.M
Quant Title : GC Extractables
QLast Update : Tue Mar 01 06:48:49 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

