

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD031221\
 Data File : PD061577.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Mar 2021 21:35
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Ankita
 3/15/2021 10:51:50 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 13 01:02:42 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031021.M
 Quant Title : GC Extractables
 QLast Update : Thu Mar 11 04:34:41 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.486	3.983	42777601	378.6E6	53.367	45.385
28)	SA Decachlor...	8.261	9.041	55079360	397.9E6	45.673	46.112
Target Compounds							
2)	A alpha-BHC	3.926	4.372	67605233	599.7E6	54.860	46.268
3)	MA gamma-BHC...	4.212	4.656	62963924	552.1E6	53.494	45.505
4)	MA Heptachlor	4.515	5.164	57132483	489.7E6	46.051	38.439
5)	MB Aldrin	4.765	5.469	63126291	551.0E6	53.892	46.075
6)	B beta-BHC	4.462	4.825	26835299	221.1E6	50.754	45.345
7)	B delta-BHC	4.668	5.042	65893594	550.4E6	55.762	45.893
8)	B Heptachlo...	5.209	5.855	58352439	537.5E6	51.338	49.387
9)	A Endosulfan I	5.548	6.212	55778140	475.5E6	51.342m	45.703
10)	B gamma-Chl...	5.437	6.087	61031171	504.3E6	52.443	44.590m
11)	B alpha-Chl...	5.496	6.161	59820605	504.6E6	51.679	46.285
12)	B 4,4'-DDE	5.659	6.318	61733337	486.2E6	56.217	43.779
13)	MA Dieldrin	5.794	6.468	59952177	493.1E6	52.240	44.624
14)	MA Endrin	6.052	6.685	55118091	424.2E6	52.189	45.250
15)	B Endosulfa...	6.326	6.894	50125065	396.9E6	50.155	44.390
16)	A 4,4'-DDD	6.176	6.807	49934800	374.1E6	53.742	41.780
17)	MA 4,4'-DDT	6.418	7.105	47959001	394.4E6	48.232	42.641
18)	B Endrin al...	6.495	7.018	39541454	319.1E6	47.876	44.416
19)	B Endosulfa...	6.709	7.241	51898580	410.5E6	49.921	44.571
20)	A Methoxychlor	6.964	7.562	25438478	187.5E6	45.129	40.524m
21)	B Endrin ke...	7.202	7.712	60832857	419.6E6	49.236	45.488
22)	Mirex	7.393	8.166	42004313	259.9E6	42.366	42.791

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD031221\
 Data File : PD061577.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Mar 2021 21:35
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Ankita
 3/15/2021 10:51:50 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 13 01:02:42 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031021.M
 Quant Title : GC Extractables
 QLast Update : Thu Mar 11 04:34:41 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

