

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD031322\
 Data File : PD068529.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Mar 2022 03:16
 Operator : AR\AJ
 Sample : INDB301
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDB3033

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 14 06:52:37 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031322CLP.M
 Quant Title : GC Extractables
 QLast Update : Mon Mar 14 06:52:20 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.449	4.176	26378421	215.0E6	20.000	20.000
27)	SA Decachlor...	8.212	9.376	54641403	251.2E6	40.000	40.000
Target Compounds							
5)	MB Aldrin	4.719	5.704	34682477	269.6E6	20.000	20.000
6)	B beta-BHC	4.423	5.026	16559003	128.5E6	20.000	20.000
7)	B delta-BHC	4.627	5.252	34820949	281.6E6	20.000	20.000
8)	B Heptachlo...	5.164	6.092	32002011	240.1E6	20.000	20.000
10)	B trans-Chl...	5.390	6.329	32797238	237.5E6	20.000	20.000
11)	B cis-Chlor...	5.449	6.404	32311277	229.5E6	20.000	20.000
12)	B 4,4'-DDE	5.615	6.554	59093486	408.0E6	40.000	40.000
15)	B Endosulfa...	6.279	7.139	57592427	358.5E6	40.000	40.000
18)	B Endrin al...	6.449	7.263	43554278	266.2E6	40.000	40.000
19)	B Endosulfa...	6.661	7.488	55387304	327.1E6	40.000	40.000
21)	B Endrin ke...	7.155	7.964	63620587	341.2E6	40.000	40.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD031322\
Data File : PD068529.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Mar 2022 03:16
Operator : AR\AJ
Sample : INDB301
Misc :
ALS Vial : 34 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 14 06:52:37 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031322CLP.M
Quant Title : GC Extractables
QLast Update : Mon Mar 14 06:52:20 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

