

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD031323\
Data File : PD074177.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Mar 2023 13:57
Operator : AR\AJ
Sample : INDB101
Misc :
ALS Vial : 10 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
INDB1026

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 03/14/2023
Supervised By :Ankita Jodhani 03/14/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 14 00:51:48 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031323CLP.M
Quant Title : GC Extractables
QLast Update : Tue Mar 14 00:51:41 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(7) delta-BHC (B)

4.784min 5.555 ng/ml

response 19857747

(7) delta-BHC #2 (B)

4.132min 5.045 ng/ml

response 8868274

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD031323\
Data File : PD074177.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Mar 2023 13:57
Operator : AR\AJ
Sample : INDB101
Misc :
ALS Vial : 10 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
INDB1026

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 03/14/2023
Supervised By :Ankita Jodhani 03/14/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 14 00:51:48 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031323CLP.M
Quant Title : GC Extractables
QLast Update : Tue Mar 14 00:51:41 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(7) delta-BHC (B)

4.783min 4.990 ng/ml m

response 17838223

(7) delta-BHC #2 (B)

4.132min 5.045 ng/ml

response 8868274

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD031323\
Data File : PD074177.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Mar 2023 13:57
Operator : AR\AJ
Sample : INDB101
Misc :
ALS Vial : 10 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
INDB1026

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 03/14/2023
Supervised By :Ankita Jodhani 03/14/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 14 00:51:48 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031323CLP.M
Quant Title : GC Extractables
QLast Update : Tue Mar 14 00:51:41 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(8) Heptachlor epoxide (B)

5.704min 5.384 ng/ml

response 18206722

(8) Heptachlor epoxide #2 (B)

4.731min 5.157 ng/ml

response 8313018

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD031323\
Data File : PD074177.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Mar 2023 13:57
Operator : AR\AJ
Sample : INDB101
Misc :
ALS Vial : 10 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
INDB1026

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 03/14/2023
Supervised By :Ankita Jodhani 03/14/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 14 00:51:48 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031323CLP.M
Quant Title : GC Extractables
QLast Update : Tue Mar 14 00:51:41 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(8) Heptachlor epoxide (B)

5.702min 5.287 ng/ml m

response 17880774

(8) Heptachlor epoxide #2 (B)

4.731min 5.157 ng/ml

response 8313018