

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD031422\
 Data File : PD0684570.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Mar 2022 19:04
 Operator : AR\AJ
 Sample : N1820-03
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 15 07:19:35 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031322CLP.M
 Quant Title : GC Extractables
 QLast Update : Mon Mar 14 07:24:04 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.447	4.173	27686391	214.6E6	20.331	19.607
27)	SA Decachlor...	8.206	9.365	53441596	236.7E6	38.432	37.535
Target Compounds							
2)	A alpha-BHC	0.000	4.598f	0	411069	N.D.	<MDL
3)	MA gamma-BHC...	0.000	4.861	0	927654	N.D.	<MDL
6)	B beta-BHC	0.000	5.004	0	8739999	N.D.	1.368
7)	B delta-BHC	0.000	5.225f	0	4732065	N.D.	0.334
8)	B Heptachlo...	0.000	6.144f	0	1239032	N.D.	0.103
10)	B trans-Chl...	0.000	6.326	0	1416471	N.D.	0.119
11)	B cis-Chlor...	0.000	6.357f	0	2347176	N.D.	0.205
13)	MA Dieldrin	0.000	6.723	0	1146906	N.D.	0.103
14)	MA Endrin	0.000	6.997f	0	470561	N.D.	<MDL
15)	B Endosulfa...	0.000	7.133	0	5786493	N.D.	0.647
16)	A 4,4'-DDD	0.000	6.997f	0	470561	N.D.	<MDL
19)	B Endosulfa...	0.000	7.544f	0	207547	N.D.	<MDL
21)	B Endrin ke...	0.000	7.942	0	4360949	N.D.	0.511
25)	Toxaphene-4	6.587	7.544f	1520900	207547	54.668	1.553 #
26)	Toxaphene-5	0.000	7.942f	0	4360949	N.D.	49.769


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD031422\
Data File : PD0684570.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Mar 2022 19:04
Operator : AR\AJ
Sample : N1820-03
Misc :
ALS Vial : 34 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 15 07:19:35 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031322CLP.M
Quant Title : GC Extractables
QLast Update : Mon Mar 14 07:24:04 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 3.447 min
Delta R.T.: -0.002 min
Response: 27686391
Conc: 20.33 ng/ml

Instrument :
ECD_D
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.173 min
Delta R.T.: -0.003 min
Response: 214611014
Conc: 19.61 ng/ml




#6 beta-BHC

R.T.: 0.000 min
Exp R.T. : 4.422 min
Response: 0
Conc: N.D.



#6 beta-BHC


R.T.: 5.004 min
Delta R.T.: -0.021 min
Response: 8739999
Conc: 1.37 ng/ml



#7 delta-BHC


R.T.: 0.000 min
Exp R.T.: 4.625 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :




#7 delta-BHC

R.T.: 5.225 min
Delta R.T.: -0.026 min
Response: 4732065
Conc: 0.33 ng/ml




#8 Heptachlor epoxide

R.T.: 0.000 min
Exp R.T.: 5.163 min
Response: 0
Conc: N.D.



#8 Heptachlor epoxide


R.T.: 6.144 min
Delta R.T.: 0.052 min
Response: 1239032
Conc: 0.10 ng/ml



#10 trans-Chlordane


R.T.: 0.000 min
Exp R.T.: 5.389 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :




#10 trans-Chlordane

R.T.: 6.326 min
Delta R.T.: -0.002 min
Response: 1416471
Conc: 0.12 ng/ml




#11 cis-Chlordane

R.T.: 0.000 min
Exp R.T.: 5.448 min
Response: 0
Conc: N.D.



#11 cis-Chlordane


R.T.: 6.357 min
Delta R.T.: -0.046 min
Response: 2347176
Conc: 0.20 ng/ml



#13 Dieldrin


R.T.: 0.000 min
Exp R.T.: 5.746 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :




#13 Dieldrin

R.T.: 6.723 min
Delta R.T.: 0.010 min
Response: 1146906
Conc: 0.10 ng/ml




#15 Endosulfan II

R.T.: 0.000 min
Exp R.T.: 6.279 min
Response: 0
Conc: N.D.



#15 Endosulfan II


R.T.: 7.133 min
Delta R.T.: -0.005 min
Response: 5786493
Conc: 0.65 ng/ml



#21 Endrin ketone


R.T.: 0.000 min
Exp R.T.: 7.155 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :




#21 Endrin ketone

R.T.: 7.942 min
Delta R.T.: -0.021 min
Response: 4360949
Conc: 0.51 ng/ml




#25 Toxaphene-4

R.T.: 6.587 min
Delta R.T.: 0.029 min
Response: 1520900
Conc: 54.67 ng/ml



#25 Toxaphene-4


R.T.: 7.544 min
Delta R.T.: 0.065 min
Response: 207547
Conc: 1.55 ng/ml



#26 Toxaphene-5


R.T.: 0.000 min
Exp R.T.: 7.045 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :




#26 Toxaphene-5

R.T.: 7.942 min
Delta R.T.: 0.051 min
Response: 4360949
Conc: 49.77 ng/ml



#27 Decachlorobiphenyl

R.T.: 8.206 min
Delta R.T.: -0.006 min
Response: 53441596
Conc: 38.43 ng/ml



#27 Decachlorobiphenyl

R.T.: 9.365 min
Delta R.T.: -0.010 min
Response: 236742115
Conc: 37.54 ng/ml