

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD031422\
 Data File : PD068567.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Mar 2022 18:22
 Operator : AR\AJ
 Sample : PB143279BS
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PLCS279

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 15 07:18:56 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031322CLP.M
 Quant Title : GC Extractables
 QLast Update : Mon Mar 14 07:24:04 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.448	4.172	29908067	223.5E6	21.963	20.421
27)	SA Decachlor...	8.206	9.366	64602205	281.2E6	46.458	44.591
Target Compounds							
2)	A alpha-BHC	3.887	4.568	9550508	79016310	4.989	5.054
3)	MA gamma-BHC...	4.170	4.859	9027756	75234271	4.973	5.189
4)	MA Heptachlor	4.469	5.386	9167629	73612199	5.118	5.348
5)	MB Aldrin	4.716	5.698	8709040	66397695	4.952	4.920
6)	B beta-BHC	4.421	5.022	4749640	35304382	5.700	5.528
7)	B delta-BHC	4.624	5.247	7763926	72742130	4.391	5.139
8)	B Heptachlo...	5.160	6.087	9056271	72603711	5.576	6.028
9)	A Endosulfan I	5.499	6.450	8493966	59984327	5.517	5.512
10)	B trans-Chl...	5.386	6.323	9334085	65737305	5.593	5.527
11)	B cis-Chlor...	5.445	6.398	9335055	65453419	5.687	5.707
12)	B 4,4'-DDE	5.612	6.548	16438445	116.6E6	11.051	11.416
13)	MA Dieldrin	5.743	6.708	17539907	123.6E6	10.500	11.122
14)	MA Endrin	6.001	6.927	14918090	83368753	10.552	9.861
15)	B Endosulfa...	6.275	7.133	16500865	104.8E6	11.376	11.715
16)	A 4,4'-DDD	6.128	7.039	14484826	97446401	11.020	11.614
17)	MA 4,4'-DDT	6.369	7.342	14306268	91603248	10.713	11.623
18)	B Endrin al...	6.444	7.257	11890626	74537795	10.843	11.225
19)	B Endosulfa...	6.657	7.482	16234075	96451332	11.616	11.776
20)	A Methoxychlor	6.920	7.795	42128852	213.7E6	59.519	61.000
21)	B Endrin ke...	7.150	7.957	19524171	106.9E6	12.236	12.518

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD031422\
Data File : PD068567.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Mar 2022 18:22
Operator : AR\AJ
Sample : PB143279BS
Misc :
ALS Vial : 31 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 15 07:18:56 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031322CLP.M
Quant Title : GC Extractables
QLast Update : Mon Mar 14 07:24:04 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

