

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD031521\
 Data File : PD061586.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Mar 2021 17:20
 Operator : AR\AJ
 Sample : PSTDICC100
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDICC100

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 16 02:22:30 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031521.M
 Quant Title : GC Extractables
 QLast Update : Tue Mar 16 02:09:18 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.486	3.982	77295816	751.3E6	98.266	97.446
28)	SA Decachlor...	8.258	9.039	106.3E6	783.3E6	96.540	95.579
Target Compounds							
2)	A alpha-BHC	3.926	4.370	124.8E6	1159.4E6	100.236	98.004
3)	MA gamma-BHC...	4.210	4.655	112.9E6	1036.0E6	99.821	97.615
4)	MA Heptachlor	4.514	5.163	80123685	948.7E6	98.673	97.443
5)	MB Aldrin	4.764	5.468	113.7E6	1061.6E6	99.629	97.373
6)	B beta-BHC	4.460	4.824	47543492	421.1E6	96.362	97.060
7)	B delta-BHC	4.667	5.041	116.6E6	1065.0E6	99.601	97.991
8)	B Heptachlo...	5.208	5.854	105.6E6	948.9E6	99.188	97.692
9)	A Endosulfan I	5.548	6.210	101.5E6	913.7E6	98.879	97.236
10)	B gamma-Chl...	5.436	6.087	108.5E6	1003.2E6	99.081	97.450
11)	B alpha-Chl...	5.494	6.160	106.7E6	955.5E6	98.848	97.277
12)	B 4,4'-DDE	5.657	6.316	108.0E6	985.9E6	99.639	97.218
13)	MA Dieldrin	5.793	6.467	105.4E6	984.0E6	99.248	97.696
14)	MA Endrin	6.051	6.684	92538173	817.8E6	98.702	97.498
15)	B Endosulfa...	6.324	6.893	92762093	798.6E6	97.932	96.821
16)	A 4,4'-DDD	6.174	6.805	78906054	777.3E6	98.858	97.179
17)	MA 4,4'-DDT	6.416	7.104	82292091	840.0E6	98.871	97.720
18)	B Endrin al...	6.493	7.017	72907889	653.4E6	97.196	96.980
19)	B Endosulfa...	6.706	7.240	95394028	849.0E6	97.947	96.708
20)	A Methoxychlor	6.963	7.561	39576432	364.1E6	97.158	97.281
21)	B Endrin ke...	7.200	7.710	114.2E6	855.2E6	97.653	96.194
22)	Mirex	7.391	8.164	81156550	546.8E6	96.595	94.773

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD031521\
Data File : PD061586.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Mar 2021 17:20
Operator : AR\AJ
Sample : PSTDICC100
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDICC100

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 16 02:22:30 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031521.M
Quant Title : GC Extractables
QLast Update : Tue Mar 16 02:09:18 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

