

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD031621\
 Data File : PD061643.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Mar 2021 14:01
 Operator : AR\AJ
 Sample : PTOXCCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PTOXCCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 17 05:48:57 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX031521.M
 Quant Title : GC Extractables
 QLast Update : Tue Mar 16 03:01:54 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds						
1) SA Tetrachlo...	3.487	3.981	38717038	384.2E6	50.155	51.477
7) SA Decachlor...	8.259	9.039	54425490	401.5E6	49.216	49.285
Target Compounds						
2) Toxaphene-1	5.971	6.556	5502804	44716877	530.314	476.141
3) Toxaphene-2	6.239	7.063	4206603	37134814	517.951	468.463
4) Toxaphene-3	6.604	7.139	9343706	107.0E6	536.389	484.991
5) Toxaphene-4	6.749	7.225	3233495	67572300	535.393	490.420
6) Toxaphene-5	7.031	7.639	4090524	49120914	529.747	502.724


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD031621\
Data File : PD061643.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Mar 2021 14:01
Operator : AR\AJ
Sample : PTOXCCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 17 05:48:57 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX031521.M
Quant Title : GC Extractables
QLast Update : Tue Mar 16 03:01:54 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 3.487 min
Delta R.T.: 0.001 min
Response: 38717038
Conc: 50.16 ng/ml

Instrument :
ECD_D
ClientSampleId :
PTOXCCC500




#1 Tetrachloro-m-xylene

R.T.: 3.981 min
Delta R.T.: 0.000 min
Response: 384214473
Conc: 51.48 ng/ml




#2 Toxaphene-1

R.T.: 5.971 min
Delta R.T.: 0.002 min
Response: 5502804
Conc: 530.31 ng/ml



#2 Toxaphene-1


R.T.: 6.556 min
Delta R.T.: 0.000 min
Response: 44716877
Conc: 476.14 ng/ml



#3 Toxaphene-2


R.T.: 6.239 min
Delta R.T.: 0.003 min
Response: 4206603
Conc: 517.95 ng/ml

Instrument :
ECD_D
ClientSampleId :
PTOXCCC500




#3 Toxaphene-2

R.T.: 7.063 min
Delta R.T.: 0.000 min
Response: 37134814
Conc: 468.46 ng/ml




#4 Toxaphene-3

R.T.: 6.604 min
Delta R.T.: 0.002 min
Response: 9343706
Conc: 536.39 ng/ml



#4 Toxaphene-3


R.T.: 7.139 min
Delta R.T.: 0.000 min
Response: 106988155
Conc: 484.99 ng/ml



#5 Toxaphene-4


R.T.: 6.749 min
Delta R.T.: 0.001 min
Response: 3233495
Conc: 535.39 ng/ml

Instrument :
ECD_D
ClientSampleId :
PTOXCCC500




#5 Toxaphene-4

R.T.: 7.225 min
Delta R.T.: 0.000 min
Response: 67572300
Conc: 490.42 ng/ml




#6 Toxaphene-5

R.T.: 7.031 min
Delta R.T.: -0.002 min
Response: 4090524
Conc: 529.75 ng/ml



#6 Toxaphene-5


R.T.: 7.639 min
Delta R.T.: 0.000 min
Response: 49120914
Conc: 502.72 ng/ml



#7 Decachlorobiphenyl

R.T.: 8.259 min
Delta R.T.: 0.000 min
Response: 54425490
Conc: 49.22 ng/ml

Instrument :
ECD_D
ClientSampleId :
PTOXCCC500



#7 Decachlorobiphenyl

R.T.: 9.039 min
Delta R.T.: 0.000 min
Response: 401495579
Conc: 49.29 ng/ml