

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD031623\
 Data File : PD074259.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Mar 2023 14:16
 Operator : AR\AJ
 Sample : 01884-12
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 20 21:38:36 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031323CLP.M
 Quant Title : GC Extractables
 QLast Update : Tue Mar 14 03:05:06 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.551	2.772	37772353	18956730	15.204	15.068
27)	SA Decachlor...	9.098	7.926	73409009	35211100	26.821	29.406
Target Compounds							
5)	MB Aldrin	0.000	4.207	0	267157	N.D.	0.156
6)	B beta-BHC	4.554	3.863f	779102	390768	0.461	0.482
7)	B delta-BHC	4.779	0.000	4068382	0	1.118	N.D. #
8)	B Heptachlo...	5.698	4.687f	1953534	16772	0.595	<MDL #
9)	A Endosulfan I	6.099	0.000	431649	0	0.145	N.D. #
10)	B trans-Chl...	6.025f	0.000	312437	0	<MDL	N.D. #
11)	B cis-Chlor...	6.025	0.000	312437	0	<MDL	N.D. #
13)	MA Dieldrin	6.372	0.000	142418	0	<MDL	N.D. #
15)	B Endosulfa...	6.800	5.941	334266	94669	0.121	<MDL #
16)	A 4,4'-DDD	6.727	5.875f	141291	190251	<MDL	0.162 #
17)	MA 4,4'-DDT	7.077	0.000	251052	0	0.100	N.D. #
19)	B Endosulfa...	7.173	0.000	385255	0	0.148	N.D. #
20)	A Methoxychlor	7.519	0.000	131008	0	<MDL	N.D. #
21)	B Endrin ke...	7.700	0.000	270140	0	<MDL	N.D. #
24)	Toxaphene-3	6.727f	5.941	141291	94669	6.714	8.918 #
25)	Toxaphene-4	7.173	0.000	385255	0	6.220	N.D. #
26)	Toxaphene-5	7.604	0.000	207923	0	4.570	N.D. #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD031623\
Data File : PD074259.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Mar 2023 14:16
Operator : AR\AJ
Sample : 01884-12
Misc :
ALS Vial : 15 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 20 21:38:36 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031323CLP.M
Quant Title : GC Extractables
QLast Update : Tue Mar 14 03:05:06 2023
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.551 min
Delta R.T.: 0.000 min
Response: 37772353
Conc: 15.20 ng/ml




#1 Tetrachloro-m-xylene

R.T.: 2.772 min
Delta R.T.: -0.017 min
Response: 18956730
Conc: 15.07 ng/ml




#5 Aldrin

R.T.: 0.000 min
Exp R.T.: 5.278 min
Response: 0
Conc: N.D.




#5 Aldrin

R.T.: 4.207 min
Delta R.T.: -0.022 min
Response: 267157
Conc: 0.16 ng/ml




#6 beta-BHC

R.T.: 4.554 min
Delta R.T.: 0.018 min
Response: 779102
Conc: 0.46 ng/ml




#6 beta-BHC

R.T.: 3.863 min
Delta R.T.: -0.038 min
Response: 390768
Conc: 0.48 ng/ml




#7 delta-BHC

R.T.: 4.779 min
Delta R.T.: -0.006 min
Response: 4068382
Conc: 1.12 ng/ml




#7 delta-BHC

R.T.: 0.000 min
Exp R.T.: 4.133 min
Response: 0
Conc: N.D.




#8 Heptachlor epoxide

R.T.: 5.698 min
Delta R.T.: -0.007 min
Response: 1953534
Conc: 0.60 ng/ml




#8 Heptachlor epoxide

R.T.: 4.687 min
Delta R.T.: -0.044 min
Response: 16772
Conc: N.D.




#9 Endosulfan I

R.T.: 6.099 min
Delta R.T.: 0.008 min
Response: 431649
Conc: 0.14 ng/ml




#9 Endosulfan I

R.T.: 0.000 min
Exp R.T.: 5.118 min
Response: 0
Conc: N.D.




#15 Endosulfan II

R.T.: 6.800 min
Delta R.T.: -0.017 min
Response: 334266
Conc: 0.12 ng/ml




#15 Endosulfan II

R.T.: 5.941 min
Delta R.T.: 0.004 min
Response: 94669
Conc: N.D.




#16 4,4'-DDD

R.T.: 6.727 min
Delta R.T.: -0.002 min
Response: 141291
Conc: N.D.




#16 4,4'-DDD

R.T.: 5.875 min
Delta R.T.: 0.067 min
Response: 190251
Conc: 0.16 ng/ml




#17 4,4'-DDT

R.T.: 7.077 min
Delta R.T.: 0.032 min
Response: 251052
Conc: 0.10 ng/ml




#17 4,4'-DDT

R.T.: 0.000 min
Exp R.T. : 6.059 min
Response: 0
Conc: N.D.




#19 Endosulfan Sulfate

R.T.: 7.173 min
Delta R.T.: -0.009 min
Response: 385255
Conc: 0.15 ng/ml




#19 Endosulfan Sulfate

R.T.: 0.000 min
Exp R.T. : 6.339 min
Response: 0
Conc: N.D.




#24 Toxaphene-3

R.T.: 6.727 min
Delta R.T.: 0.065 min
Response: 141291
Conc: 6.71 ng/ml




#24 Toxaphene-3

R.T.: 5.941 min
Delta R.T.: -0.032 min
Response: 94669
Conc: 8.92 ng/ml




#25 Toxaphene-4

R.T.: 7.173 min
Delta R.T.: 0.000 min
Response: 385255
Conc: 6.22 ng/ml




#25 Toxaphene-4

R.T.: 0.000 min
Exp R.T.: 6.608 min
Response: 0
Conc: N.D.




#26 Toxaphene-5

R.T.: 7.604 min
Delta R.T.: 0.011 min
Response: 207923
Conc: 4.57 ng/ml




#26 Toxaphene-5

R.T.: 0.000 min
Exp R.T.: 6.732 min
Response: 0
Conc: N.D.



#27 Decachlorobiphenyl

R.T.: 9.098 min
Delta R.T.: 0.000 min
Response: 73409009
Conc: 26.82 ng/ml



#27 Decachlorobiphenyl

R.T.: 7.926 min
Delta R.T.: -0.016 min
Response: 35211100
Conc: 29.41 ng/ml