

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD031721\
 Data File : PD061670.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Mar 2021 16:40
 Operator : AR\AJ
 Sample : PCHLORICC750
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 17 18:02:04 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031721.M
 Quant Title : GC Extractables
 QLast Update : Wed Mar 17 18:01:27 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.485	3.981	75246998	628.6E6	74.831	73.297
28)	SA Decachlor...	8.257	9.039	87098850	620.3E6	74.184	73.105
Target Compounds							
23)	Chlordane-1	4.360	4.974	24574804	277.7E6	736.520	727.673
24)	Chlordane-2	4.863	5.441	28910208	271.2E6	717.382	708.301
25)	Chlordane-3	5.435	6.087	91190782	1195.1E6	751.906	738.398
26)	Chlordane-4	5.493	6.161	74798761	1373.3E6	750.741	734.534
27)	Chlordane-5	6.333	6.965	34414180	261.2E6	751.523	746.982



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD031721\
Data File : PD061670.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Mar 2021 16:40
Operator : AR\AJ
Sample : PCHLORICC750
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 17 18:02:04 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031721.M
Quant Title : GC Extractables
QLast Update : Wed Mar 17 18:01:27 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 3.485 min
Delta R.T.: 0.000 min
Response: 75246998
Conc: 74.83 ng/ml

Instrument :
ECD_D
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.981 min
Delta R.T.: 0.000 min
Response: 628636280
Conc: 73.30 ng/ml




#23 Chlordane-1

R.T.: 4.360 min
Delta R.T.: 0.000 min
Response: 24574804
Conc: 736.52 ng/ml



#23 Chlordane-1


R.T.: 4.974 min
Delta R.T.: 0.000 min
Response: 277658131
Conc: 727.67 ng/ml



#24 Chlordane-2


R.T.: 4.863 min
Delta R.T.: 0.000 min
Response: 28910208
Conc: 717.38 ng/ml

Instrument :
ECD_D
ClientSampleId :




#24 Chlordane-2

R.T.: 5.441 min
Delta R.T.: 0.000 min
Response: 271157173
Conc: 708.30 ng/ml




#25 Chlordane-3

R.T.: 5.435 min
Delta R.T.: 0.000 min
Response: 91190782
Conc: 751.91 ng/ml



#25 Chlordane-3


R.T.: 6.087 min
Delta R.T.: 0.000 min
Response: 1195093530
Conc: 738.40 ng/ml



#26 Chlordane-4


R.T.: 5.493 min
Delta R.T.: 0.000 min
Response: 74798761
Conc: 750.74 ng/ml

Instrument :
ECD_D
ClientSampleId :




#26 Chlordane-4

R.T.: 6.161 min
Delta R.T.: 0.000 min
Response: 1373251637
Conc: 734.53 ng/ml




#27 Chlordane-5

R.T.: 6.333 min
Delta R.T.: 0.000 min
Response: 34414180
Conc: 751.52 ng/ml



#27 Chlordane-5


R.T.: 6.965 min
Delta R.T.: 0.000 min
Response: 261169079
Conc: 746.98 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.257 min
Delta R.T.: 0.000 min
Response: 87098850
Conc: 74.18 ng/ml

Instrument :
ECD_D
ClientSampleId :



#28 Decachlorobiphenyl

R.T.: 9.039 min
Delta R.T.: 0.000 min
Response: 620271127
Conc: 73.10 ng/ml