Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD031723\
Data File : PD074280.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Mar 2023 18:29
Operator : AR\AJ
Sample : 01627-08
Misc : TOX LOQ 100PPB WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
LOQ-WATER-02-QT1-2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 20 11:08:04 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX031423.M
Quant Title : GC Extractables
QLast Update : Wed Mar 15 05:13:38 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0. Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1) SA	Tetrachlo...	3.550	2.771	44347125	22399322	19.741	19.287
7) SA	Decachlor...	9.100	7.927	54346027	24301790	20.767	19.937
	Target Compounds						
2)	Toxaphene-1	6.259	5.008	2580629	963790	104.686	103.797
3)	Toxaphene-2	7.174	5.692	5889068	807068	97.388	84.473
4)	Toxaphene-3	7.272	6.609	3412138	2494517	98.285	79.253
5)	Toxaphene-4	7.596	6.735	4137488	3305393	95.124	79.116
6)	Toxaphene-5	7.961	7.051	4537870	1463589	97.962	82.738

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD031723\
Data File : PD074280.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Mar 2023 18:29
Operator : AR\AJ
Sample : 01627-08
Misc : TOX LOQ 100PPB WATER
ALS Vial : 14 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 20 11:08:04 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX031423.M
Quant Title : GC Extractables
QLast Update : Wed Mar 15 05:13:38 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0. Signal #2 Info : 30M x 0.32mm x0.25µm