

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Data\PD031819\  
 Data File : PD051840.D  
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH  
 Acq On : 18 Mar 2019 14:32  
 Operator : AJ\SJ  
 Sample : K1930-04  
 Misc :  
 ALS Vial : 22 Sample Multiplier: 1

Instrument :  
 ECD\_D  
 ClientSampleId :  
 TP-3

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Mar 19 02:33:42 2019  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Method\PD030219.M  
 Quant Title : GC Extractables  
 QLast Update : Fri Mar 01 23:43:17 2019  
 Response via : Initial Calibration  
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl  
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1  
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

|                             | Compound        | RT#1  | RT#2  | Resp#1   | Resp#2  | ng/ml  | ng/ml  |
|-----------------------------|-----------------|-------|-------|----------|---------|--------|--------|
| -----                       |                 |       |       |          |         |        |        |
| System Monitoring Compounds |                 |       |       |          |         |        |        |
| 1)                          | SA Tetrachlo... | 3.223 | 4.044 | 27949643 | 547.9E6 | 21.079 | 19.107 |
| 28)                         | SA Decachlor... | 7.822 | 9.155 | 19095905 | 302.1E6 | 16.800 | 16.535 |

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Data\PD031819\  
Data File : PD051840.D  
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH  
Acq On : 18 Mar 2019 14:32  
Operator : AJ\SJ  
Sample : K1930-04  
Misc :  
ALS Vial : 22 Sample Multiplier: 1

Instrument :  
ECD\_D  
ClientSampleId :  
TP-3

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Mar 19 02:33:42 2019  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Method\PD030219.M  
Quant Title : GC Extractables  
QLast Update : Fri Mar 01 23:43:17 2019  
Response via : Initial Calibration  
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl  
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1  
Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 3.223 min  
Delta R.T.: 0.000 min  
Response: 27949643  
Conc: 21.08 ng/ml

Instrument :  
ECD\_D  
ClientSampleId :  
TP-3




#1 Tetrachloro-m-xylene

R.T.: 4.044 min  
Delta R.T.: 0.000 min  
Response: 547924140  
Conc: 19.11 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.822 min  
Delta R.T.: -0.002 min  
Response: 19095905  
Conc: 16.80 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.155 min  
Delta R.T.: 0.000 min  
Response: 302052051  
Conc: 16.54 ng/ml