

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD031821\
 Data File : PD061711.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Mar 2021 11:45
 Operator : AR\AJ
 Sample : M1691-01
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 FK-GF-02-068-A

Manual Integrations
 APPROVED

Ankita
 3/19/2021 10:23:55 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 19 06:29:52 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031721.M
 Quant Title : GC Extractables
 QLast Update : Thu Mar 18 07:52:17 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.485	3.982	12106164	121.5E6	14.163	14.164
28)	SA Decachlor...	8.259	9.040	17240904	126.0E6	14.088	14.492
Target Compounds							
4)	MA Heptachlor	4.513	5.164	1993843	15945842	1.693	1.318
10)	B gamma-Chl...	5.436	6.087	6158896	120.1E6	5.188m	10.639 #
11)	B alpha-Chl...	5.494	6.162	16654988	316.9E6	14.210	29.018 #
12)	B 4,4'-DDE	5.656	6.315	5418770	68660116	4.774	6.190 #
16)	A 4,4'-DDD	6.175	6.810	1293740	24407092	1.351m	2.699m#
17)	MA 4,4'-DDT	6.416	7.105	14740619	129.7E6	14.665	13.157

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD031821\
Data File : PD061711.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2021 11:45
Operator : AR\AJ
Sample : M1691-01
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
FK-GF-02-068-A

Manual Integrations
APPROVED

Ankita
3/19/2021 10:23:55 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 06:29:52 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031721.M
Quant Title : GC Extractables
QLast Update : Thu Mar 18 07:52:17 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

