

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD031922\
 Data File : PD068689.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Mar 2022 12:46
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 19 00:45:19 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031222.M
 Quant Title : GC Extractables
 QLast Update : Mon Mar 14 03:30:32 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.448	4.175	61033363	476.0E6	50.267	49.784
28)	SA Decachlor...	8.207	9.370	61317288	278.8E6	48.534	49.207
Target Compounds							
2)	A alpha-BHC	3.888	4.571	94022691	702.6E6	51.930	51.053
3)	MA gamma-BHC...	4.171	4.861	86582630	635.9E6	51.853	50.861
4)	MA Heptachlor	4.469	5.389	84734524	611.0E6	51.433	51.411
5)	MB Aldrin	4.717	5.702	83531263	590.9E6	52.268	51.482
6)	B beta-BHC	4.421	5.024	37598453	277.1E6	50.498	50.482
7)	B delta-BHC	4.625	5.250	85657168	619.1E6	52.415	51.485
8)	B Heptachlo...	5.161	6.090	76905893	526.6E6	51.425	50.428
9)	A Endosulfan I	5.500	6.453	72093133	477.3E6	51.523	50.894
10)	B gamma-Chl...	5.387	6.327	78009395	522.2E6	51.425	50.662
11)	B alpha-Chl...	5.446	6.401	76658678	505.2E6	51.289	50.743
12)	B 4,4'-DDE	5.612	6.551	71504567	475.1E6	52.267	51.273
13)	MA Dieldrin	5.744	6.711	75653188	490.7E6	51.812	50.515
14)	MA Endrin	6.002	6.931	62117590	368.5E6	47.155	47.630
15)	B Endosulfa...	6.276	7.136	63507976	381.0E6	50.955	50.412
16)	A 4,4'-DDD	6.129	7.042	61493737	382.5E6	52.147	51.797
17)	MA 4,4'-DDT	6.370	7.345	59700654	367.9E6	49.669	50.078
18)	B Endrin al...	6.445	7.260	50631603	307.1E6	51.001	50.748
19)	B Endosulfa...	6.658	7.485	64731393	369.5E6	50.618	50.237
20)	A Methoxychlor	6.920	7.798	33119916	178.8E6	47.790	49.043
21)	B Endrin ke...	7.151	7.960	76847109	394.8E6	52.200	50.944
22)	Mirex	7.335	8.426	57202506	269.6E6	48.265	49.187

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD031922\
Data File : PD068689.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2022 12:46
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 00:45:19 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031222.M
Quant Title : GC Extractables
QLast Update : Mon Mar 14 03:30:32 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

