

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD032023\
 Data File : PD074301.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Mar 2023 17:07
 Operator : AR\AJ
 Sample : PTOXCCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PTOXCCC500

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 03/21/2023
 Supervised By : Ankita Jodhani 03/21/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 21 04:41:49 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX031423.M
 Quant Title : GC Extractables
 QLast Update : Wed Mar 15 05:13:38 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x 0. Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds						
1) SA Tetrachlo...	3.550	2.771	104.2E6	56474439	46.366	48.628
7) SA Decachlor...	9.094	7.924	126.0E6	58095327	48.164	47.661
Target Compounds						
2) Toxaphene-1	6.257	5.005	12009617	4361914	487.181	469.763
3) Toxaphene-2	7.172	5.691	29762181	4869651	492.178	509.688
4) Toxaphene-3	7.269	6.607	16723017	16069144	481.699	510.529
5) Toxaphene-4	7.591	6.732	21847881	20783377	502.297m	497.458
6) Toxaphene-5	7.958	7.049	23141197	9038879	499.565	510.978


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD032023\
Data File : PD074301.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Mar 2023 17:07
Operator : AR\AJ
Sample : PTOXCCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 21 04:41:49 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX031423.M
Quant Title : GC Extractables
QLast Update : Wed Mar 15 05:13:38 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0. Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.550 min
Delta R.T.: 0.000 min
Response: 104157681
Conc: 46.37 ng/ml

Instrument :
ECD_D
ClientSampleId :
PTOXCCC500


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 03/21/2023
Supervised By :Ankita Jodhani 03/21/2023




#1 Tetrachloro-m-xylene

R.T.: 2.771 min
Delta R.T.: 0.000 min
Response: 56474439
Conc: 48.63 ng/ml




#2 Toxaphene-1

R.T.: 6.257 min
Delta R.T.: -0.003 min
Response: 12009617
Conc: 487.18 ng/ml



#2 Toxaphene-1

R.T.: 5.005 min
Delta R.T.: -0.002 min
Response: 4361914
Conc: 469.76 ng/ml




#3 Toxaphene-2

R.T.: 7.172 min
Delta R.T.: -0.003 min
Response: 29762181
Conc: 492.18 ng/ml

Instrument :
ECD_D
ClientSampleId :
PTOXCCC500


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 03/21/2023
Supervised By :Ankita Jodhani 03/21/2023




#3 Toxaphene-2

R.T.: 5.691 min
Delta R.T.: -0.002 min
Response: 4869651
Conc: 509.69 ng/ml




#4 Toxaphene-3

R.T.: 7.269 min
Delta R.T.: -0.003 min
Response: 16723017
Conc: 481.70 ng/ml



#4 Toxaphene-3

R.T.: 6.607 min
Delta R.T.: -0.003 min
Response: 16069144
Conc: 510.53 ng/ml




#5 Toxaphene-4

R.T.: 7.591 min
Delta R.T.: -0.005 min
Response: 21847881
Conc: 502.30 ng/ml

Instrument :
ECD_D
ClientSampleId :
PTOXCCC500


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 03/21/2023
Supervised By :Ankita Jodhani 03/21/2023




#5 Toxaphene-4

R.T.: 6.732 min
Delta R.T.: -0.003 min
Response: 20783377
Conc: 497.46 ng/ml




#6 Toxaphene-5

R.T.: 7.958 min
Delta R.T.: -0.003 min
Response: 23141197
Conc: 499.57 ng/ml



#6 Toxaphene-5

R.T.: 7.049 min
Delta R.T.: -0.002 min
Response: 9038879
Conc: 510.98 ng/ml




#7 Decachlorobiphenyl

R.T.: 9.094 min
Delta R.T.: -0.006 min
Response: 126038218
Conc: 48.16 ng/ml

Instrument :
ECD_D
ClientSampleId :
PTOXCCC500

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 03/21/2023
Supervised By :Ankita Jodhani 03/21/2023



#7 Decachlorobiphenyl

R.T.: 7.924 min
Delta R.T.: -0.003 min
Response: 58095327
Conc: 47.66 ng/ml