

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD032119\
 Data File : PD051930.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 21 Mar 2019 15:48
 Operator : AJ\SJ
 Sample : TOXAPH401
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 TOXAPH401

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 22 08:38:26 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD032119CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 22 08:37:11 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.222	4.044	58468889	1250.2E6	39.086	36.229
2)	SA Decachlor...	7.822	9.156	92322874	1509.1E6	77.626	65.345
Target Compounds							
22)	Toxaphene-1	5.522	6.447	54252861	601.2E6	3864.555	3501.222
23)	Toxaphene-2	5.585	6.638	92134096	1122.2E6	3901.449	3566.451
24)	Toxaphene-3	5.954	7.148	159.2E6	1003.6E6	4090.891	3697.449
25)	Toxaphene-4	6.194	7.224	159.7E6	2697.0E6	4125.558	3710.577
26)	Toxaphene-5	6.667	7.726	183.3E6	1101.2E6	4046.115	3612.938

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD032119\
Data File : PD051930.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 21 Mar 2019 15:48
Operator : AJ\SJ
Sample : TOXAPH401
Misc :
ALS Vial : 7 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 22 08:38:26 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD032119CLP.M
Quant Title : GC Extractables
QLast Update : Fri Mar 22 08:37:11 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

