

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD032122\
 Data File : PD068707.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Mar 2022 09:58
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 22 01:13:20 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031222.M
 Quant Title : GC Extractables
 QLast Update : Mon Mar 14 03:30:32 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.457	4.176	61540102	486.7E6	50.685	50.910
2)	SA Decachlor...	8.218	9.376	58857942	266.7E6	46.588	47.072
Target Compounds							
2)	A alpha-BHC	3.897	4.572	95307543	713.3E6	52.639	51.827
3)	MA gamma-BHC...	4.181	4.863	87351843	636.9E6	52.314	50.936
4)	MA Heptachlor	4.480	5.391	83200206	596.2E6	50.502	50.171
5)	MB Aldrin	4.728	5.703	84071338	587.3E6	52.606	51.168
6)	B beta-BHC	4.431	5.026	37661209	274.8E6	50.582	50.054
7)	B delta-BHC	4.636	5.252	85824608	605.0E6	52.517	50.306
8)	B Heptachlo...	5.172	6.092	76963324	517.5E6	51.464	49.558
9)	A Endosulfan I	5.511	6.455	72363468	471.1E6	51.716	50.228
10)	B gamma-Chl...	5.398	6.329	77985306	512.9E6	51.410	49.764
11)	B alpha-Chl...	5.457	6.404	76659477	496.2E6	51.289	49.841
12)	B 4,4'-DDE	5.623	6.553	71372223	464.9E6	52.170	50.171
13)	MA Dieldrin	5.755	6.713	75598699	481.8E6	51.775	49.598
14)	MA Endrin	6.013	6.934	62761364	361.3E6	47.643	46.691
15)	B Endosulfa...	6.287	7.139	63409284	371.7E6	50.876	49.169
16)	A 4,4'-DDD	6.141	7.045	62397682	377.2E6	52.914	51.069
17)	MA 4,4'-DDT	6.382	7.348	56137506	346.0E6	46.705	47.093
18)	B Endrin al...	6.457	7.263	47629089	289.7E6	47.977	47.868
19)	B Endosulfa...	6.670	7.488	64125607	363.6E6	50.144	49.437
20)	A Methoxychlor	6.932	7.801	30919523	164.5E6	44.615	45.125
21)	B Endrin ke...	7.163	7.963	75366744	386.0E6	51.195	49.816
22)	Mirex	7.347	8.429	56300382	262.6E6	47.504	47.907

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD031222\
Data File : PD068707.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Mar 2022 09:58
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 22 01:13:20 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031222.M
Quant Title : GC Extractables
QLast Update : Mon Mar 14 03:30:32 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

