

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD032423\
Data File : PD074454.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Mar 2023 10:30
Operator : AR\AJ
Sample : PEM030
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_D
LabSampleId :
PEM030

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 03/27/2023
Supervised By :Ankita Jodhani 03/27/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 24 14:28:20 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031323CLP.M
Quant Title : GC Extractables
QLast Update : Tue Mar 14 03:05:06 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(12) 4,4'-DDE (B)
6.208min 0.347 ng/ml
response 1038794

(12) 4,4'-DDE #2 (B)
5.186min 0.014 ng/ml
response 20605

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD032423\
Data File : PD074454.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Mar 2023 10:30
Operator : AR\AJ
Sample : PEM030
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :

ECD_D

LabSampleId :

PEM030


Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 03/27/2023

Supervised By :Ankita Jodhani 03/27/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 24 14:28:20 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031323CLP.M
Quant Title : GC Extractables
QLast Update : Tue Mar 14 03:05:06 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(12) 4,4'-DDE (B)

6.208min 0.347 ng/ml

response 1038794

(12) 4,4'-DDE #2 (B)

5.234min 0.179 ng/ml m

response 260622

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD032423\
Data File : PD074454.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Mar 2023 10:30
Operator : AR\AJ
Sample : PEM030
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :

ECD_D

LabSampleId :

PEM030



Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 03/27/2023

Supervised By :Ankita Jodhani 03/27/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 24 14:28:20 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031323CLP.M
Quant Title : GC Extractables
QLast Update : Tue Mar 14 03:05:06 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(14) Endrin (MA)

6.595min 48.616 ng/ml

response 128116226

(14) Endrin #2 (MA)

5.641min 47.717 ng/ml

response 64193154

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD032423\
Data File : PD074454.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Mar 2023 10:30
Operator : AR\AJ
Sample : PEM030
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :

ECD_D

LabSampleId :

PEM030



Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 03/27/2023

Supervised By :Ankita Jodhani 03/27/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 24 14:28:20 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031323CLP.M
Quant Title : GC Extractables
QLast Update : Tue Mar 14 03:05:06 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(14) Endrin (MA)

6.594min 46.561 ng/ml m

response 122701699

(14) Endrin #2 (MA)

5.641min 47.717 ng/ml

response 64193154

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD032423\
Data File : PD074454.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Mar 2023 10:30
Operator : AR\AJ
Sample : PEM030
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :

ECD_D

LabSampleId :

PEM030


Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 03/27/2023

Supervised By :Ankita Jodhani 03/27/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 24 14:28:20 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031323CLP.M
Quant Title : GC Extractables
QLast Update : Tue Mar 14 03:05:06 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(21) Endrin ketone (B)

7.666min 3.706 ng/ml

response 10988283

(21) Endrin ketone #2 (B)

6.844min 3.887 ng/ml

response 5914917

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD032423\
Data File : PD074454.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Mar 2023 10:30
Operator : AR\AJ
Sample : PEM030
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_D
LabSampleId :
PEM030

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 03/27/2023
Supervised By :Ankita Jodhani 03/27/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 24 14:28:20 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031323CLP.M
Quant Title : GC Extractables
QLast Update : Tue Mar 14 03:05:06 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(21) Endrin ketone (B)
7.665min 3.766 ng/ml m
response 11166735

(21) Endrin ketone #2 (B)
6.844min 3.887 ng/ml
response 5914917