

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD032522\
 Data File : PD068760.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Mar 2022 14:32
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 23 16:44:25 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031222.M
 Quant Title : GC Extractables
 QLast Update : Mon Mar 14 03:30:32 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.446	4.171	60380668	494.7E6	49.730	51.739
28)	SA Decachlor...	8.204	9.365	61604778	295.5E6	48.762	52.146
Target Compounds							
2)	A alpha-BHC	3.885	4.567	92290507	727.7E6	50.973	52.876
3)	MA gamma-BHC...	4.169	4.857	84626933	653.9E6	50.682	52.300
4)	MA Heptachlor	4.467	5.384	80360339	613.5E6	48.778	51.624
5)	MB Aldrin	4.715	5.697	80655225	606.1E6	50.469	52.802
6)	B beta-BHC	4.419	5.020	36785297	286.4E6	49.405	52.170
7)	B delta-BHC	4.623	5.246	83653960	634.8E6	51.189	52.783
8)	B Heptachlo...	5.159	6.085	74482280	539.9E6	49.805	51.704
9)	A Endosulfan I	5.497	6.448	69870325	492.0E6	49.934	52.461
10)	B gamma-Chl...	5.385	6.322	75654489	537.3E6	49.873	52.128
11)	B alpha-Chl...	5.444	6.397	74307195	520.4E6	49.715	52.265
12)	B 4,4'-DDE	5.610	6.547	69262242	488.5E6	50.628	52.717
13)	MA Dieldrin	5.741	6.707	73060929	507.6E6	50.037	52.257
14)	MA Endrin	5.999	6.926	55567592	356.4E6	42.182	46.059
15)	B Endosulfa...	6.273	7.132	61887343	395.2E6	49.655	52.284
16)	A 4,4'-DDD	6.127	7.038	60413259	403.7E6	51.231	54.658
17)	MA 4,4'-DDT	6.367	7.341	55589403	375.1E6	46.249	51.058
18)	B Endrin al...	6.443	7.255	51018572	330.9E6	51.391	54.684
19)	B Endosulfa...	6.656	7.480	63383131	388.3E6	49.564	52.789
20)	A Methoxychlor	6.918	7.794	31237675	185.6E6	45.074	50.915
21)	B Endrin ke...	7.148	7.955	77909100	427.5E6	52.922	55.176
22)	Mirex	7.333	8.421	55359089	286.7E6	46.710	52.302

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD032522\
Data File : PD068760.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Mar 2022 14:32
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 23 16:44:25 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031222.M
Quant Title : GC Extractables
QLast Update : Mon Mar 14 03:30:32 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

