

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD032719\
 Data File : PD052219.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 27 Mar 2019 21:12
 Operator : AJ\SJ
 Sample : K2085-15
 Misc :
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 BF7TO

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 28 03:16:38 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD032519CLP.M
 Quant Title : GC Extractables
 QLast Update : Mon Mar 25 18:59:03 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.221	4.042	21615842	445.8E6	15.229	15.320
27)	SA Decachlor...	7.816	9.152	38524219	434.3E6	24.602	28.134
Target Compounds							
2)	A alpha-BHC	0.000	4.406f	0	165.4E6	N.D.	3.782
3)	MA gamma-BHC...	3.921	4.680f	6655410	30043641	2.970	0.756 #
4)	MA Heptachlor	0.000	5.258	0	16664270	N.D.	0.415
6)	B beta-BHC	0.000	4.934f	0	39575541	N.D.	2.157
7)	B delta-BHC	0.000	5.080f	0	58502788	N.D.	1.474
8)	B Heptachlo...	0.000	5.907	0	104.0E6	N.D.	3.053
9)	A Endosulfan I	0.000	6.261	0	51132908	N.D.	1.631
10)	B trans-Chl...	0.000	6.146	0	196.4E6	N.D.	5.783
11)	B cis-Chlor...	0.000	6.261	0	51132908	N.D.	1.575
17)	MA 4,4'-DDT	0.000	7.199	0	12764975	N.D.	0.661
24)	Toxaphene-3	0.000	7.199	0	12764975	N.D.	63.134
25)	Toxaphene-4	0.000	7.244	0	23011230	N.D.	41.343


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD032719\
Data File : PD052219.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 27 Mar 2019 21:12
Operator : AJ\SJ
Sample : K2085-15
Misc :
ALS Vial : 46 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 28 03:16:38 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD032519CLP.M
Quant Title : GC Extractables
QLast Update : Mon Mar 25 18:59:03 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 3.221 min
Delta R.T.: 0.000 min
Response: 21615842
Conc: 15.23 ng/ml

Instrument :
ECD_D
ClientSampleId :
BF7TO




#1 Tetrachloro-m-xylene

R.T.: 4.042 min
Delta R.T.: -0.001 min
Response: 445776383
Conc: 15.32 ng/ml




#2 alpha-BHC

R.T.: 0.000 min
Exp R.T. : 3.641 min
Response: 0
Conc: N.D.



#2 alpha-BHC


R.T.: 4.406 min
Delta R.T.: -0.029 min
Response: 165377174
Conc: 3.78 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 3.921 min
Delta R.T.: 0.011 min
Response: 6655410
Conc: 2.97 ng/ml

Instrument :
ECD_D
ClientSampleId :
BF7TO




#3 gamma-BHC (Lindane)

R.T.: 4.680 min
Delta R.T.: -0.042 min
Response: 30043641
Conc: 0.76 ng/ml




#4 Heptachlor

R.T.: 0.000 min
Exp R.T. : 4.187 min
Response: 0
Conc: N.D.



#4 Heptachlor


R.T.: 5.258 min
Delta R.T.: 0.023 min
Response: 16664270
Conc: 0.42 ng/ml



#6 beta-BHC


R.T.: 0.000 min
Exp R.T.: 4.154 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
BF7TO




#6 beta-BHC

R.T.: 4.934 min
Delta R.T.: 0.041 min
Response: 39575541
Conc: 2.16 ng/ml




#7 delta-BHC

R.T.: 0.000 min
Exp R.T.: 4.344 min
Response: 0
Conc: N.D.



#7 delta-BHC


R.T.: 5.080 min
Delta R.T.: -0.032 min
Response: 58502788
Conc: 1.47 ng/ml



#8 Heptachlor epoxide


R.T.: 0.000 min
Exp R.T.: 4.846 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
BF7TO




#8 Heptachlor epoxide

R.T.: 5.907 min
Delta R.T.: -0.024 min
Response: 104018943
Conc: 3.05 ng/ml




#9 Endosulfan I

R.T.: 0.000 min
Exp R.T.: 5.168 min
Response: 0
Conc: N.D.



#9 Endosulfan I


R.T.: 6.261 min
Delta R.T.: -0.029 min
Response: 51132908
Conc: 1.63 ng/ml



#10 trans-Chlordane


R.T.: 0.000 min
Exp R.T.: 5.063 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
BF7TO




#10 trans-Chlordane

R.T.: 6.146 min
Delta R.T.: -0.020 min
Response: 196403711
Conc: 5.78 ng/ml




#11 cis-Chlordane

R.T.: 0.000 min
Exp R.T.: 5.119 min
Response: 0
Conc: N.D.



#11 cis-Chlordane


R.T.: 6.261 min
Delta R.T.: 0.022 min
Response: 51132908
Conc: 1.57 ng/ml



#17 4,4'-DDT


R.T.: 0.000 min
Exp R.T.: 6.021 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
BF7TO




#17 4,4'-DDT

R.T.: 7.199 min
Delta R.T.: 0.012 min
Response: 12764975
Conc: 0.66 ng/ml




#24 Toxaphene-3

R.T.: 0.000 min
Exp R.T.: 5.954 min
Response: 0
Conc: N.D.



#24 Toxaphene-3


R.T.: 7.199 min
Delta R.T.: 0.052 min
Response: 12764975
Conc: 63.13 ng/ml



#25 Toxaphene-4


R.T.: 0.000 min
Exp R.T.: 6.192 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
BF7TO




#25 Toxaphene-4

R.T.: 7.244 min
Delta R.T.: 0.022 min
Response: 23011230
Conc: 41.34 ng/ml



#27 Decachlorobiphenyl

R.T.: 7.816 min
Delta R.T.: -0.003 min
Response: 38524219
Conc: 24.60 ng/ml



#27 Decachlorobiphenyl

R.T.: 9.152 min
Delta R.T.: -0.004 min
Response: 434297423
Conc: 28.13 ng/ml