

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD032719\
 Data File : PD052225.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 27 Mar 2019 22:37
 Operator : AJ\SJ
 Sample : K2084-13
 Misc :
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 BF7X3

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 28 03:17:26 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD032519CLP.M
 Quant Title : GC Extractables
 QLast Update : Mon Mar 25 18:59:03 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.221	4.042	20174023	390.9E6	14.213	13.435
27)	SA Decachlor...	7.816	9.152	32529135	373.2E6	20.774	24.175
Target Compounds							
2)	A alpha-BHC	0.000	4.405f	0	140.7E6	N.D.	3.218
3)	MA gamma-BHC...	0.000	4.684f	0	30030976	N.D.	0.756
4)	MA Heptachlor	0.000	5.197f	0	198.9E6	N.D.	4.959
7)	B delta-BHC	0.000	5.080f	0	76353432	N.D.	1.924
8)	B Heptachlo...	0.000	5.906	0	98691368	N.D.	2.897
9)	A Endosulfan I	0.000	6.244f	0	45809883	N.D.	1.461
10)	B trans-Chl...	0.000	6.181	0	68220473	N.D.	2.009
11)	B cis-Chlor...	0.000	6.244	0	45809883	N.D.	1.411
15)	B Endosulfa...	0.000	6.979	0	11622174	N.D.	0.458
18)	B Endrin al...	0.000	7.082	0	19553632	N.D.	0.955
24)	Toxaphene-3	0.000	7.082	0	19553632	N.D.	96.709


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD032719\
Data File : PD052225.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 27 Mar 2019 22:37
Operator : AJ\SJ
Sample : K2084-13
Misc :
ALS Vial : 49 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 28 03:17:26 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD032519CLP.M
Quant Title : GC Extractables
QLast Update : Mon Mar 25 18:59:03 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 3.221 min
Delta R.T.: 0.000 min
Response: 20174023
Conc: 14.21 ng/ml

Instrument :
ECD_D
ClientSampleId :
BF7X3




#1 Tetrachloro-m-xylene

R.T.: 4.042 min
Delta R.T.: -0.001 min
Response: 390911629
Conc: 13.43 ng/ml




#2 alpha-BHC

R.T.: 0.000 min
Exp R.T. : 3.641 min
Response: 0
Conc: N.D.



#2 alpha-BHC


R.T.: 4.405 min
Delta R.T.: -0.030 min
Response: 140714856
Conc: 3.22 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 0.000 min
Exp R.T.: 3.910 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
BF7X3




#3 gamma-BHC (Lindane)

R.T.: 4.684 min
Delta R.T.: -0.038 min
Response: 30030976
Conc: 0.76 ng/ml




#4 Heptachlor

R.T.: 0.000 min
Exp R.T.: 4.187 min
Response: 0
Conc: N.D.



#4 Heptachlor


R.T.: 5.197 min
Delta R.T.: -0.039 min
Response: 198900493
Conc: 4.96 ng/ml



#7 delta-BHC


R.T.: 0.000 min
Exp R.T.: 4.344 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
BF7X3




#7 delta-BHC

R.T.: 5.080 min
Delta R.T.: -0.032 min
Response: 76353432
Conc: 1.92 ng/ml




#8 Heptachlor epoxide

R.T.: 0.000 min
Exp R.T.: 4.846 min
Response: 0
Conc: N.D.



#8 Heptachlor epoxide


R.T.: 5.906 min
Delta R.T.: -0.024 min
Response: 98691368
Conc: 2.90 ng/ml



#9 Endosulfan I


R.T.: 0.000 min
Exp R.T.: 5.168 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
BF7X3




#9 Endosulfan I

R.T.: 6.244 min
Delta R.T.: -0.045 min
Response: 45809883
Conc: 1.46 ng/ml




#10 trans-Chlordane

R.T.: 0.000 min
Exp R.T.: 5.063 min
Response: 0
Conc: N.D.



#10 trans-Chlordane


R.T.: 6.181 min
Delta R.T.: 0.015 min
Response: 68220473
Conc: 2.01 ng/ml



#11 cis-Chlordane


R.T.: 0.000 min
Exp R.T.: 5.119 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
BF7X3




#11 cis-Chlordane

R.T.: 6.244 min
Delta R.T.: 0.006 min
Response: 45809883
Conc: 1.41 ng/ml




#15 Endosulfan II

R.T.: 0.000 min
Exp R.T.: 5.920 min
Response: 0
Conc: N.D.



#15 Endosulfan II


R.T.: 6.979 min
Delta R.T.: 0.003 min
Response: 11622174
Conc: 0.46 ng/ml



#18 Endrin aldehyde


R.T.: 0.000 min
Exp R.T.: 6.087 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
BF7X3




#18 Endrin aldehyde

R.T.: 7.082 min
Delta R.T.: -0.018 min
Response: 19553632
Conc: 0.95 ng/ml




#24 Toxaphene-3

R.T.: 0.000 min
Exp R.T.: 5.954 min
Response: 0
Conc: N.D.



#24 Toxaphene-3


R.T.: 7.082 min
Delta R.T.: -0.065 min
Response: 19553632
Conc: 96.71 ng/ml



#27 Decachlorobiphenyl

R.T.: 7.816 min
Delta R.T.: -0.003 min
Response: 32529135
Conc: 20.77 ng/ml

Instrument :
ECD_D
ClientSampleId :
BF7X3



#27 Decachlorobiphenyl

R.T.: 9.152 min
Delta R.T.: -0.004 min
Response: 373188427
Conc: 24.18 ng/ml