

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD032720\
 Data File : PD057665.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Mar 2020 18:50
 Operator : AJ\MA
 Sample : PB127350BS
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PB127350BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 28 04:15:26 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD032720.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 27 14:08:11 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.227	3.897	14337762	15616176	20.081	19.905
28)	SA Decachlor...	7.866	8.899	15289022	18841754	20.081	21.423
Target Compounds							
23)	Chlordane-1	4.056	4.876	4622871	5864830	185.730	191.340
24)	Chlordane-2	4.541	5.337	5515810	8252812	192.397	222.629
25)	Chlordane-3	5.090	5.978	15971329	28185865	181.947	193.471
26)	Chlordane-4	5.145	6.052	15006445	34539561	185.565	195.439
27)	Chlordane-5	5.962	6.850	5111204	6840717	191.746	205.490



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD032720\
Data File : PD057665.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2020 18:50
Operator : AJ\MA
Sample : PB127350BS
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PB127350BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 28 04:15:26 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD032720.M
Quant Title : GC Extractables
QLast Update : Fri Mar 27 14:08:11 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 3.227 min
Delta R.T.: 0.000 min
Response: 14337762
Conc: 20.08 ng/ml

Instrument :
ECD_D
ClientSampleId :
PB127350BS




#1 Tetrachloro-m-xylene

R.T.: 3.897 min
Delta R.T.: 0.000 min
Response: 15616176
Conc: 19.90 ng/ml




#23 Chlordane-1

R.T.: 4.056 min
Delta R.T.: 0.002 min
Response: 4622871
Conc: 185.73 ng/ml



#23 Chlordane-1


R.T.: 4.876 min
Delta R.T.: 0.001 min
Response: 5864830
Conc: 191.34 ng/ml



#24 Chlordane-2


R.T.: 4.541 min
Delta R.T.: 0.002 min
Response: 5515810
Conc: 192.40 ng/ml

Instrument :
ECD_D
ClientSampleId :
PB127350BS




#24 Chlordane-2

R.T.: 5.337 min
Delta R.T.: 0.002 min
Response: 8252812
Conc: 222.63 ng/ml




#25 Chlordane-3

R.T.: 5.090 min
Delta R.T.: 0.000 min
Response: 15971329
Conc: 181.95 ng/ml



#25 Chlordane-3


R.T.: 5.978 min
Delta R.T.: 0.000 min
Response: 28185865
Conc: 193.47 ng/ml



#26 Chlordane-4


R.T.: 5.145 min
Delta R.T.: 0.000 min
Response: 15006445
Conc: 185.57 ng/ml

Instrument :
ECD_D
ClientSampleId :
PB127350BS




#26 Chlordane-4

R.T.: 6.052 min
Delta R.T.: 0.000 min
Response: 34539561
Conc: 195.44 ng/ml




#27 Chlordane-5

R.T.: 5.962 min
Delta R.T.: 0.000 min
Response: 5111204
Conc: 191.75 ng/ml



#27 Chlordane-5


R.T.: 6.850 min
Delta R.T.: 0.000 min
Response: 6840717
Conc: 205.49 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.866 min
Delta R.T.: 0.000 min
Response: 15289022
Conc: 20.08 ng/ml

Instrument :
ECD_D
ClientSampleId :
PB127350BS



#28 Decachlorobiphenyl

R.T.: 8.899 min
Delta R.T.: 0.000 min
Response: 18841754
Conc: 21.42 ng/ml