

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD032823\
 Data File : PD074535.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Mar 2023 01:04
 Operator : AR\AJ
 Sample : 02061-03MSD
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 CB961MSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 29 04:09:28 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031323CLP.M
 Quant Title : GC Extractables
 QLast Update : Tue Mar 14 03:05:06 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.549	2.771	44579980	21315661	17.944	16.943
27)	SA Decachlor...	9.094	7.921	72166051	27688818	26.367	23.124
Target Compounds							
2)	A alpha-BHC	4.006	3.274	170.9E6	77258334	43.789	43.231
3)	MA gamma-BHC...	4.340	3.604	157.7E6	68546438	42.725	40.132
4)	MA Heptachlor	4.932	3.945	154.5E6	66052594	40.810	39.837
5)	MB Aldrin	5.275	4.226	157.9E6	67306309	43.206	39.311
6)	B beta-BHC	4.534	3.900	66586202	30252488	39.387	37.341
7)	B delta-BHC	4.783	4.130	147.6E6	71138499	40.567	40.220
8)	B Heptachlo...	5.702	4.729	137.4E6	61098979	41.854	38.568
9)	A Endosulfan I	6.089	5.099	130.3E6	56564236	43.679	40.340
10)	B trans-Chl...	5.958	4.979	138.2E6	61160239	43.067	39.534
11)	B cis-Chlor...	6.038	5.044	139.7E6	61097136	42.631	38.624
12)	B 4,4'-DDE	6.210	5.233	272.4E6	117.0E6	90.930	80.540
13)	MA Dieldrin	6.365	5.365	311.2E6	133.1E6	98.094	85.860
14)	MA Endrin	6.595	5.640	252.1E6	112.1E6	95.671	83.339
15)	B Endosulfa...	6.814	5.933	246.3E6	104.3E6	89.218	77.563
16)	A 4,4'-DDD	6.728	5.788	223.4E6	96733389	94.534	82.158
17)	MA 4,4'-DDT	7.043	6.040	241.4E6	103.7E6	96.209	82.815
18)	B Endrin al...	6.944	6.113	178.4E6	75976967	79.186	69.115
19)	B Endosulfa...	7.179	6.336	232.5E6	97609377	89.270	77.542
20)	A Methoxychlor	7.519	6.616	652.7E6	249.8E6	475.445	406.102
21)	B Endrin ke...	7.666	6.843	270.7E6	122.7E6	91.298	80.657

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD032823\
Data File : PD074535.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Mar 2023 01:04
Operator : AR\AJ
Sample : 02061-03MSD
Misc :
ALS Vial : 34 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 29 04:09:28 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD031323CLP.M
Quant Title : GC Extractables
QLast Update : Tue Mar 14 03:05:06 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

